

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121420\
 Data File : VN065113.D
 Acq On : 14 Dec 2020 20:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

MMDadoda
 12/16/2020 1:41:11 PM

Quant Time: Dec 15 05:53:31 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121420W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 15 05:41:47 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	333516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	471411	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.379	117	446180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	223624	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	189472	49.41	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 98.82%			
35) Dibromofluoromethane	7.547	113	166144	50.70	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.40%			
50) Toluene-d8	10.058	98	580999	50.14	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.28%			
62) 4-Bromofluorobenzene	12.376	95	205646	51.53	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 103.06%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.833	85	158249	47.98	ug/l	100
3) Chloromethane	2.036	50	172093	59.53	ug/l	99
4) Vinyl Chloride	2.164	62	147327	47.60	ug/l	99
5) Bromomethane	2.528	94	86058	54.03	ug/l	96
6) Chloroethane	2.669	64	78746	47.15	ug/l	97
7) Trichlorofluoromethane	2.984	101	280003	49.97	ug/l	99
8) Diethyl Ether	3.373	74	95631	49.81	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.721	101	152460	46.26	ug/l	95
10) Methyl Iodide	3.914	142	198073	43.22	ug/l	98
11) Tert butyl alcohol	4.733	59	116608	285.42	ug/l	98
12) 1,1-Dichloroethene	3.698	96	152838	49.77	ug/l	94
13) Acrolein	3.566	56	70621	378.77	ug/l	100
14) Allyl chloride	4.277	41	248560	50.55	ug/l	95
15) Acrylonitrile	4.930	53	372231	274.86	ug/l	99
16) Acetone	3.772	43	298780	209.23	ug/l	100
17) Carbon Disulfide	4.013	76	413752	48.60	ug/l	99
18) Methyl Acetate	4.277	43	166954	54.00	ug/l	94
19) Methyl tert-butyl Ether	4.991	73	476630	51.56	ug/l	98
20) Methylene Chloride	4.505	84	175755	51.76	ug/l	96
21) trans-1,2-Dichloroethene	4.991	96	166248	51.08	ug/l	99
22) Diisopropyl ether	5.901	45	526641	53.75	ug/l	96
23) Vinyl Acetate	5.843	43	2175012	273.52	ug/l	98
24) 1,1-Dichloroethane	5.798	63	297991	49.96	ug/l	98
25) 2-Butanone	6.785	43	484048	264.87	ug/l	99
26) 2,2-Dichloropropane	6.775	77	235303	44.70	ug/l	99
27) cis-1,2-Dichloroethene	6.782	96	186677	51.37	ug/l	99
28) Bromochloromethane	7.151	49	127104	48.48	ug/l	99
29) Tetrahydrofuran	7.167	42	324111	288.80	ug/l	97
30) Chloroform	7.328	83	308847	48.75	ug/l	99
31) Cyclohexane	7.614	56	249969	47.99	ug/l	95
32) 1,1,1-Trichloroethane	7.531	97	280736	49.55	ug/l	99
36) 1,1-Dichloropropene	7.753	75	221574	49.22	ug/l	96
37) Ethyl Acetate	6.878	43	207591	53.99	ug/l	97
38) Carbon Tetrachloride	7.733	117	248878	47.92	ug/l	98
39) Methylcyclohexane	9.045	83	248172	50.74	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121420\
 Data File : VN065113.D
 Acq On : 14 Dec 2020 20:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

MMDadoda
 12/16/2020 1:41:11 PM

Quant Time: Dec 15 05:53:31 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121420W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Dec 15 05:41:47 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
40) Benzene	8.003	78	673968	51.22	ug/l	99
41) Methacrylonitrile	7.126	41	110801	60.88	ug/l #	75
42) 1,2-Dichloroethane	8.087	62	238998	50.19	ug/l	99
43) Isopropyl Acetate	8.126	43	339821	53.69	ug/l	99
44) Trichloroethene	8.801	130	193085	49.43	ug/l	95
45) 1,2-Dichloropropane	9.084	63	178009	50.88	ug/l	98
46) Dibromomethane	9.174	93	120464	50.64	ug/l	93
47) Bromodichloromethane	9.370	83	244841	50.26	ug/l	99
48) Methyl methacrylate	9.164	41	165147	54.39	ug/l	94
49) 1,4-Dioxane	9.164	88	55345	1224.35	ug/l	96
51) 4-Methyl-2-Pentanone	9.952	43	1043210	284.94	ug/l	95
52) Toluene	10.122	92	431994	53.42	ug/l	98
53) t-1,3-Dichloropropene	10.351	75	256384	51.35	ug/l	99
54) cis-1,3-Dichloropropene	9.807	75	280116	52.66	ug/l	97
55) 1,1,2-Trichloroethane	10.531	97	169732	51.94	ug/l	96
56) Ethyl methacrylate	10.399	69	231850	50.42	ug/l	93
57) 1,3-Dichloropropane	10.675	76	281862	52.97	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.662	63	582371	274.97	ug/l	100
59) 2-Hexanone	10.720	43	744665	293.42	ug/l	94
60) Dibromochloromethane	10.871	129	208066	53.71	ug/l	99
61) 1,2-Dibromoethane	10.978	107	176764	52.50	ug/l	100
64) Tetrachloroethene	10.601	164	208339	47.05	ug/l	97
65) Chlorobenzene	11.405	112	470028	50.03	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.479	131	184651	48.93	ug/l	99
67) Ethyl Benzene	11.482	91	803172	51.53	ug/l	100
68) m/p-Xylenes	11.595	106	629934	105.30	ug/l	98
69) o-Xylene	11.920	106	299186	53.51	ug/l	99
70) Styrene	11.936	104	507700	49.02	ug/l	98
71) Bromoform	12.100	173	151565	53.07	ug/l #	99
73) Isopropylbenzene	12.222	105	794643	51.45	ug/l	100
74) N-amyl acetate	12.042	43	288121	54.82	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.476	83	223077	50.07	ug/l	98
76) 1,2,3-Trichloropropane	12.527	75	212644m	52.99	ug/l	
77) Bromobenzene	12.498	156	217051	51.37	ug/l	90
78) n-propylbenzene	12.566	91	881058	52.02	ug/l	97
79) 2-Chlorotoluene	12.650	91	525363	50.83	ug/l	98
80) 1,3,5-Trimethylbenzene	12.707	105	680792	53.42	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.273	75	74528	51.68	ug/l	92
82) 4-Chlorotoluene	12.746	91	552103	51.67	ug/l	99
83) tert-Butylbenzene	12.968	119	570781	51.96	ug/l	99
84) 1,2,4-Trimethylbenzene	13.013	105	679715	52.81	ug/l	99
85) sec-Butylbenzene	13.145	105	734405	52.76	ug/l	99
86) p-Isopropyltoluene	13.264	119	690080	48.55	ug/l	99
87) 1,3-Dichlorobenzene	13.257	146	372415	51.08	ug/l	97
88) 1,4-Dichlorobenzene	13.338	146	363070	49.22	ug/l	99
89) n-Butylbenzene	13.588	91	552830	51.07	ug/l	99
90) Hexachloroethane	13.846	117	120044	51.27	ug/l	85
91) 1,2-Dichlorobenzene	13.627	146	360384	49.58	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.244	75	45447	50.59	ug/l	99
93) 1,2,4-Trichlorobenzene	14.881	180	191909	44.10	ug/l	99
94) Hexachlorobutadiene	14.981	225	114841	46.49	ug/l	99
95) Naphthalene	15.103	128	511550	45.46	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121420\
Data File : VN065113.D
Acq On : 14 Dec 2020 20:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Quant Time: Dec 15 05:53:31 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121420W.M
Quant Title : Sw846 8260
QLast Update : Tue Dec 15 05:41:47 2020
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
96) 1,2,3-Trichlorobenzene	15.283	180	178284	45.13	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Abundance Scan 1831 (7.627 min): VN065092.D\data.ms (-18) (-)





Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.627 min Scan# 1831
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Tgt Ion:168 Resp: 333516
Ion Ratio Lower Upper
168 100
99 53.0 43.0 64.4


Abundance Scan 1831 (7.627 min): VN065113.D\data.ms (-1738) (-)



Abundance Scan 29 (1.833 min): VN065092.D\data.ms (-18) (-)





Dichlorodifluoromethane
Concen: 47.98 ug/l
RT: 1.833 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion: 85 Resp: 158249
Ion Ratio Lower Upper
85 100
87 32.9 16.4 49.4


Abundance Scan 29 (1.833 min): VN065113.D\data.ms (-1) (-)



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Chloromethane
Concen: 59.53 ug/l
RT: 2.036 min Scan# 92
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 50 Resp: 172093
Ion Ratio Lower Upper
50 100
52 32.6 26.6 39.8

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Manual Integrations
APPROVED


MMDadoda
12/16/2020 1:41:11 PM



Vinyl Chloride
Concen: 47.60 ug/l
RT: 2.164 min Scan# 132
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 62 Resp: 147327
Ion Ratio Lower Upper
62 100
64 32.3 26.5 39.7









Bromomethane
 Concen: 54.03 ug/l
 RT: 2.528 min Scan# 245
 Delta R.T. -0.010 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 94 Resp: 86058
 Ion Ratio Lower Upper
 94 100
 96 97.7 75.4 113.0




Manual Integrations
 APPROVED

MMDadoda
 12/16/2020 1:41:11 PM




Chloroethane
 Concen: 47.15 ug/l
 RT: 2.669 min Scan# 289
 Delta R.T. -0.007 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion: 64 Resp: 78746
 Ion Ratio Lower Upper
 64 100
 66 33.5 25.6 38.4



Abundance Scan 389 (2.991 min): VN065092.D\data.ms (-372) #7






Trichlorofluoromethane
Concen: 49.97 ug/l
RT: 2.984 min Scan# 387
Delta R.T. -0.007 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 101 Resp: 280003
Ion Ratio Lower Upper
101 100
103 63.7 50.5 75.7


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Abundance Scan 510 (3.380 min): VN065092.D\data.ms (-492) #8



Diethyl Ether
Concen: 49.81 ug/l
RT: 3.373 min Scan# 508
Delta R.T. -0.007 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion: 74 Resp: 95631
Ion Ratio Lower Upper
74 100
45 100.2 48.9 146.8








1,1,2-Trichlorotrifluoroethane
Concen: 46.26 ug/l
RT: 3.721 min Scan# 616
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC






Tgt Ion	Ratio	Lower	Upper
101	100		
85	42.8	34.4	51.6
151	86.0	64.0	96.0




Methyl Iodide
Concen: 43.22 ug/l
RT: 3.914 min Scan# 676
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
142	100		
127	45.3	35.0	52.6
141	14.4	11.8	17.6



Abundance Scan 934 (4.743 min): VN065092.D\data.ms (-907) #11






Tert butyl alcohol
Concen: 285.42 ug/l
RT: 4.733 min Scan# 931
Delta R.T. -0.010 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 59 Resp: 116608
Ion Ratio Lower Upper
59 100
57 10.6 7.9 11.9


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 611 (3.705 min): VN065092.D\data.ms (-591) #12




1,1-Dichloroethene
Concen: 49.77 ug/l
RT: 3.698 min Scan# 609
Delta R.T. -0.007 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 96 Resp: 152838
Ion Ratio Lower Upper
96 100
61 165.6 123.4 185.2
98 62.7 49.8 74.8



Abundance Scan 570 (3.573 min): VN065092.D\data.ms (-554) #13






Acrolein
Concen: 378.77 ug/l
RT: 3.566 min Scan# 568
Delta R.T. -0.007 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 56 Resp: 70621
Ion Ratio Lower Upper
56 100
55 70.8 56.5 84.7


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 790 (4.280 min): VN065092.D\data.ms (-759) #14




Allyl chloride
Concen: 50.55 ug/l
RT: 4.277 min Scan# 789
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 41 Resp: 248560
Ion Ratio Lower Upper
41 100
39 67.9 50.9 76.3
76 33.5 29.0 43.4



Abundance Scan 993 (4.933 min): VN065092.D\data.ms (-970) #15






Acrylonitrile
 Concen: 274.86 ug/l
 RT: 4.930 min Scan# 992
 Delta R.T. -0.003 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion:	53	Resp:	372231
Ion	Ratio	Lower	Upper
53	100		
52	81.3	65.5	98.3
51	36.2	28.6	43.0


Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

MMDadoda
 12/16/2020 1:41:11 PM






Abundance Scan 634 (3.779 min): VN065092.D\data.ms (-606) #16




Acetone
 Concen: 209.23 ug/l
 RT: 3.772 min Scan# 632
 Delta R.T. -0.007 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion:	43	Resp:	298780
Ion	Ratio	Lower	Upper
43	100		
58	32.4	25.7	38.5



Abundance Scan 708 (4.017 min): VN065092.D\data.ms (-686) #17






Carbon Disulfide
Concen: 48.60 ug/l
RT: 4.013 min Scan# 707
Delta R.T. -0.004 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 76 Resp: 413752
Ion Ratio Lower Upper
76 100
78 9.5 7.4 11.0


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 790 (4.280 min): VN065092.D\data.ms (-772) #18




Methyl Acetate
Concen: 54.00 ug/l
RT: 4.277 min Scan# 789
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 43 Resp: 166954
Ion Ratio Lower Upper
43 100
74 22.3 20.2 30.2



Abundance Scan 1011 (4.991 min): VN065092.D\data.ms (-981) (-)






Methyl tert-butyl Ether
Concen: 51.56 ug/l
RT: 4.991 min Scan# 1011
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 73 Resp: 476630
Ion Ratio Lower Upper
73 100
57 22.4 17.4 26.0


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 861 (4.509 min): VN065092.D\data.ms (-839) (-)




Methylene Chloride
Concen: 51.76 ug/l
RT: 4.505 min Scan# 860
Delta R.T. -0.004 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 84 Resp: 175755
Ion Ratio Lower Upper
84 100
49 121.5 93.4 140.0
51 36.2 28.2 42.4
86 67.3 51.7 77.5



Abundance Scan 1012 (4.994 min): VN065092.D\data.ms (-98#21)






trans-1,2-Dichloroethene
Concen: 51.08 ug/l
RT: 4.991 min Scan# 1011
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	96	Resp:	166248
Ion	Ratio	Lower	Upper
96	100		
61	137.2	108.4	162.6
98	63.0	51.3	76.9


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1294 (5.901 min): VN065092.D\data.ms (-125#21)




Diisopropyl ether
Concen: 53.75 ug/l
RT: 5.901 min Scan# 1294
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	45	Resp:	526641
Ion	Ratio	Lower	Upper
45	100		
43	60.9	46.2	69.2
87	27.6	24.1	36.1
59	11.1	9.0	13.4



Abundance Scan 1276 (5.843 min): VN065092.D\data.ms (-12523)






Vinyl Acetate
Concen: 273.52 ug/l
RT: 5.843 min Scan# 1276
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 43 Resp: 2175012
Ion Ratio Lower Upper
43 100
86 10.4 8.9 13.3


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1262 (5.798 min): VN065092.D\data.ms (-12524)




1,1-Dichloroethane
Concen: 49.96 ug/l
RT: 5.798 min Scan# 1262
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 63 Resp: 297991
Ion Ratio Lower Upper
63 100
98 6.2 3.5 10.5
100 4.2 2.3 6.8



Abundance Scan 1570 (6.788 min): VN065092.D\data.ms (-15425-)



2-Butanone
Concen: 264.87 ug/l
RT: 6.785 min Scan# 1569
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion: 43 Resp: 484048
Ion Ratio Lower Upper
43 100
72 25.7 21.0 31.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM


Abundance Scan 1569 (6.785 min): VN065113.D\data.ms



Abundance Scan 1569 (6.785 min): VN065113.D\data.ms (-1477) (-)





Abundance Scan 1566 (6.775 min): VN065092.D\data.ms (-15426-)




2,2-Dichloropropane
Concen: 44.70 ug/l
RT: 6.775 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 77 Resp: 235303
Ion Ratio Lower Upper
77 100
97 24.0 11.7 35.0


Abundance Scan 1566 (6.775 min): VN065113.D\data.ms



Abundance Scan 1566 (6.775 min): VN065113.D\data.ms (-1473) (-)



Abundance Scan 1568 (6.782 min): VN065092.D\data.ms (-1547)






cis-1,2-Dichloroethene
Concen: 51.37 ug/l
RT: 6.782 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
96	100		
61	143.2	0.0	285.8
98	63.1	0.0	130.6


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1683 (7.152 min): VN065092.D\data.ms (-16628)



Bromochloromethane
Concen: 48.48 ug/l
RT: 7.151 min Scan# 1683
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.8	0.0	3.8
130	83.6	66.2	99.2



Abundance Scan 1688 (7.168 min): VN065092.D\data.ms (-16629-)

Tetrahydrofuran
Concen: 288.80 ug/l
RT: 7.167 min Scan# 1688
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion	Ratio	Lower	Upper
42	100		
72	42.6	36.6	54.8
71	40.8	33.6	50.4

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM




Abundance Scan 1688 (7.167 min): VN065113.D\data.ms (-1595-)





Abundance Scan 1738 (7.328 min): VN065092.D\data.ms (-17130-)


Chloroform
Concen: 48.75 ug/l
RT: 7.328 min Scan# 1738
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion	Ratio	Lower	Upper
83	100		
85	65.1	52.6	78.8




Abundance Scan 1738 (7.328 min): VN065113.D\data.ms (-1645-)



Abundance Scan 1828 (7.618 min): VN065092.D\data.ms (-18031-)

Cyclohexane
Concen: 47.99 ug/l
RT: 7.614 min Scan# 1827
Delta R.T. -0.004 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15






Tgt Ion:	56	Resp:	249969
Ion	Ratio	Lower	Upper
56	100		
69	35.1	27.5	41.3
84	86.7	73.9	110.9

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1800 (7.528 min): VN065092.D\data.ms (-17602-)


1,1,1-Trichloroethane
Concen: 49.55 ug/l
RT: 7.531 min Scan# 1801
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion:	97	Resp:	280736
Ion	Ratio	Lower	Upper
97	100		
99	64.0	51.7	77.5
61	42.2	33.9	50.9



Abundance Scan 1943 (7.987 min): VN065092.D\data.ms (-1925)





1,2-Dichloroethane-d4
Concen: 49.41 ug/l
RT: 7.987 min Scan# 1943
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
65	100		
67	51.7	0.0	108.2


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


MMDadoda
12/16/2020 1:41:11 PM



Abundance Scan 1943 (7.987 min): VN065113.D\data.ms (-1850)





Abundance Scan 2118 (8.550 min): VN065092.D\data.ms (-2100)




1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.550 min Scan# 2118
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion	Ratio	Lower	Upper
114	100		
63	19.5	0.0	38.6
88	15.3	0.0	32.6



Abundance Scan 2118 (8.550 min): VN065113.D\data.ms (-2025)



Abundance Scan 1807 (7.550 min): VN065092.D\data.ms (-1785)






Dibromofluoromethane
Concen: 50.70 ug/l
RT: 7.547 min Scan# 1806
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	113	Resp:	166144
Ion	Ratio	Lower	Upper
113	100		
111	99.3	81.6	122.4
192	21.7	16.2	24.2


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1870 (7.753 min): VN065092.D\data.ms (-18436)




1,1-Dichloropropene
Concen: 49.22 ug/l
RT: 7.753 min Scan# 1870
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	75	Resp:	221574
Ion	Ratio	Lower	Upper
75	100		
110	38.8	17.8	53.3
77	30.5	25.0	37.4



Abundance Scan 1599 (6.881 min): VN065092.D\data.ms (-1506) (-)






Ethyl Acetate
Concen: 53.99 ug/l
RT: 6.878 min Scan# 1598
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
43	100		
61	14.2	12.8	19.2
70	10.7	8.8	13.2


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1864 (7.734 min): VN065092.D\data.ms (-1848) (-)




Carbon Tetrachloride
Concen: 47.92 ug/l
RT: 7.733 min Scan# 1864
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
117	100		
119	98.4	76.7	115.1
121	31.4	25.4	38.2



Abundance Scan 2273 (9.049 min): VN065092.D\data.ms (-2254) (-)





Methylcyclohexane
Concen: 50.74 ug/l
RT: 9.045 min Scan# 2272
Delta R.T. -0.004 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
83	100		
55	72.6	60.2	90.2
98	45.5	37.4	56.2


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


MMDadoda
12/16/2020 1:41:11 PM



Abundance Scan 2272 (9.045 min): VN065113.D\data.ms (-2180) (-)





Abundance Scan 1948 (8.004 min): VN065092.D\data.ms (-1924) (-)




Benzene
Concen: 51.22 ug/l
RT: 8.003 min Scan# 1948
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion	Ratio	Lower	Upper
78	100		
77	23.1	17.9	26.9



Abundance Scan 1948 (8.003 min): VN065113.D\data.ms (-1855) (-)



Abundance Scan 1675 (7.126 min): VN065092.D\data.ms (-16541-)






Methacrylonitrile
Concen: 60.88 ug/l
RT: 7.126 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
41	100		
39	41.2	51.0	76.6#
67	73.4	79.8	119.8#
52	30.7	31.5	47.3#


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 1973 (8.084 min): VN065092.D\data.ms (-19542-)




1,2-Dichloroethane
Concen: 50.19 ug/l
RT: 8.087 min Scan# 1974
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
62	100		
98	9.3	0.0	19.6



Abundance Scan 1986 (8.126 min): VN065092.D\data.ms (-196743)






Isopropyl Acetate
Concen: 53.69 ug/l
RT: 8.126 min Scan# 1986
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
43	100		
61	21.0	17.2	25.8
87	13.4	11.4	17.0


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Abundance Scan 2196 (8.801 min): VN065092.D\data.ms (-21804)



Trichloroethene
Concen: 49.43 ug/l
RT: 8.801 min Scan# 2196
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
130	100		
95	94.4	0.0	197.8









1,2-Dichloropropane
 Concen: 50.88 ug/l
 RT: 9.084 min Scan# 2284
 Delta R.T. -0.000 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion: 63 Resp: 178009
 Ion Ratio Lower Upper
 63 100
 65 32.3 25.0 37.4

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC




Manual Integrations
 APPROVED

MMDadoda
 12/16/2020 1:41:11 PM




Dibromomethane
 Concen: 50.64 ug/l
 RT: 9.174 min Scan# 2312
 Delta R.T. -0.000 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion: 93 Resp: 120464
 Ion Ratio Lower Upper
 93 100
 95 83.1 68.8 103.2
 174 113.6 83.1 124.7



Abundance Scan 2373 (9.370 min): VN065092.D\data.ms (-2359) (-)







Bromodichloromethane
Concen: 50.26 ug/l
RT: 9.370 min Scan# 2373
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 83 Resp: 244841
Ion Ratio Lower Upper
83 100
85 63.6 51.8 77.8
127 8.9 6.8 10.2

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM




Methyl methacrylate
Concen: 54.39 ug/l
RT: 9.164 min Scan# 2309
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 41 Resp: 165147
Ion Ratio Lower Upper
41 100
69 85.7 71.8 107.8
39 61.2 44.1 66.1



Abundance Scan 2313 (9.177 min): VN065092.D\data.ms (-2299)






1,4-Dioxane
Concen: 1224.35 ug/l
RT: 9.164 min Scan# 2309
Delta R.T. -0.013 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	88	Resp:	55345
Ion	Ratio	Lower	Upper
88	100		
43	31.6	24.2	36.4
58	69.1	52.5	78.7


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 2587 (10.058 min): VN065092.D\data.ms (-2550)



Toluene-d8
Concen: 50.14 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion:	98	Resp:	580999
Ion	Ratio	Lower	Upper
98	100		
100	64.4	51.7	77.5



Abundance Scan 2554 (9.952 min): VN065092.D\data.ms (-2538) (-)

4-Methyl-2-Pentanone
Concen: 284.94 ug/l
RT: 9.952 min Scan# 2554
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 1043210
Ion Ratio Lower Upper
43 100
58 37.6 32.4 48.6


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 2607 (10.122 min): VN065092.D\data.ms (-2552) (-)


Toluene
Concen: 53.42 ug/l
RT: 10.122 min Scan# 2607
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion: 92 Resp: 431994
Ion Ratio Lower Upper
92 100
91 171.2 139.1 208.7



Abundance Scan 2678 (10.351 min): VN065092.D\data.ms (-2585) (-)





t-1,3-Dichloropropene
Concen: 51.35 ug/l
RT: 10.351 min Scan# 2678
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 75 Resp: 256384
Ion Ratio Lower Upper
75 100
77 32.5 25.8 38.6


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


MMDadoda
12/16/2020 1:41:11 PM



Abundance Scan 2678 (10.351 min): VN065113.D\data.ms (-2585) (-)





Abundance Scan 2509 (9.807 min): VN065092.D\data.ms (-2416) (-)




cis-1,3-Dichloropropene
Concen: 52.66 ug/l
RT: 9.807 min Scan# 2509
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion: 75 Resp: 280116
Ion Ratio Lower Upper
75 100
77 32.1 25.5 38.3
39 47.8 35.6 53.4



Abundance Scan 2509 (9.807 min): VN065113.D\data.ms (-2416) (-)



Abundance Scan 2735 (10.534 min): VN065092.D\data.ms (-2750) (-)






1,1,2-Trichloroethane
Concen: 51.94 ug/l
RT: 10.531 min Scan# 2734
Delta R.T. -0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	97	Resp:	169732
Ion	Ratio	Lower	Upper
97	100		
83	86.5	72.1	108.1
85	54.0	46.8	70.2
99	63.1	49.6	74.4


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 2693 (10.399 min): VN065092.D\data.ms (-2750) (-)




Ethyl methacrylate
Concen: 50.42 ug/l
RT: 10.399 min Scan# 2693
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	69	Resp:	231850
Ion	Ratio	Lower	Upper
69	100		
41	65.7	49.4	74.0
39	38.4	25.9	38.9



Abundance Scan 2779 (10.676 min): VN065092.D\data.ms (-2757)






1,3-Dichloropropane
Concen: 52.97 ug/l
RT: 10.675 min Scan# 2779
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 76 Resp: 281862
Ion Ratio Lower Upper
76 100
78 32.2 26.1 39.1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 2464 (9.663 min): VN065092.D\data.ms (-2459)



2-Chloroethyl Vinyl ether
Concen: 274.97 ug/l
RT: 9.662 min Scan# 2464
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 63 Resp: 582371
Ion Ratio Lower Upper
63 100
106 28.6 22.7 34.1



Abundance Scan 2793 (10.721 min): VN065092.D\data.ms (-2759) (-)

2-Hexanone
Concen: 293.42 ug/l
RT: 10.720 min Scan# 2793
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion: 43 Resp: 744665
Ion Ratio Lower Upper
43 100
58 52.5 28.3 84.9

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM

Abundance Scan 2793 (10.720 min): VN065113.D\data.ms



Abundance Scan 2793 (10.720 min): VN065113.D\data.ms (-2700) (-)





Abundance Scan 2840 (10.872 min): VN065092.D\data.ms (-2826) (-)


Dibromochloromethane
Concen: 53.71 ug/l
RT: 10.871 min Scan# 2840
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:129 Resp: 208066
Ion Ratio Lower Upper
129 100
127 77.0 39.1 117.2


Abundance Scan 2840 (10.871 min): VN065113.D\data.ms



Abundance Scan 2840 (10.871 min): VN065113.D\data.ms (-2747) (-)



Abundance Scan 2872 (10.975 min): VN065092.D\data.ms (-2653) (-)






1,2-Dibromoethane
Concen: 52.50 ug/l
RT: 10.978 min Scan# 2873
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 176764
Ion Ratio Lower Upper
107 100
109 95.0 76.0 114.0

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Abundance Scan 3307 (12.373 min): VN065092.D\data.ms (-3252) (-)



4-Bromofluorobenzene
Concen: 51.53 ug/l
RT: 12.376 min Scan# 3308
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion: 95 Resp: 205646
Ion Ratio Lower Upper
95 100
174 91.7 0.0 160.4
176 89.7 0.0 158.4





Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.379 min Scan# 2998
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Tgt Ion:	117	Resp:	446180
Ion	Ratio	Lower	Upper
117	100		
82	54.3	44.2	66.2
119	32.2	25.8	38.6

Manual Integrations
APPROVED



MMDadoda
12/16/2020 1:41:11 PM



Tetrachloroethene
Concen: 47.05 ug/l
RT: 10.601 min Scan# 2756
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Tgt Ion:	164	Resp:	208339
Ion	Ratio	Lower	Upper
164	100		
166	127.9	101.4	152.0
129	92.6	76.1	114.1
131	86.7	74.2	111.2





Abundance Scan 3006 (11.405 min): VN065092.D\data.ms (-2995) (-)


Chlorobenzene
Concen: 50.03 ug/l
RT: 11.405 min Scan# 3006
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion:112 Resp: 470028
Ion Ratio Lower Upper
112 100
114 32.2 25.1 37.7




Abundance Scan 3006 (11.405 min): VN065113.D\data.ms (-2913) (-)





Abundance Scan 3029 (11.479 min): VN065092.D\data.ms (-3068) (-)


1,1,1,2-Tetrachloroethane
Concen: 48.93 ug/l
RT: 11.479 min Scan# 3029
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion:131 Resp: 184651
Ion Ratio Lower Upper
131 100
133 95.2 47.5 142.5
119 64.5 33.3 99.8



Abundance Scan 3029 (11.479 min): VN065113.D\data.ms (-2936) (-)




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED




MMDadoda
12/16/2020 1:41:11 PM

Abundance Scan 3030 (11.483 min): VN065092.D\data.ms (-3065) (-)

Ethyl Benzene
Concen: 51.53 ug/l
RT: 11.482 min Scan# 3030
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Tgt Ion: 91 Resp: 803172
Ion Ratio Lower Upper
91 100
106 31.8 25.2 37.8






Abundance Scan 3064 (11.592 min): VN065092.D\data.ms (-3068) (-)

m/p-Xylenes
Concen: 105.30 ug/l
RT: 11.595 min Scan# 3065
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion:106 Resp: 629934
Ion Ratio Lower Upper
106 100
91 198.2 161.0 241.4




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM

Abundance Scan 3166 (11.920 min): VN065092.D\data.ms (-3166) (-)






o-Xylene
Concen: 53.51 ug/l
RT: 11.920 min Scan# 3166
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:106 Resp: 299186
Ion Ratio Lower Upper
106 100
91 208.7 105.5 316.5


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Abundance Scan 3171 (11.936 min): VN065092.D\data.ms (-3171) (-)



Styrene
Concen: 49.02 ug/l
RT: 11.936 min Scan# 3171
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:104 Resp: 507700
Ion Ratio Lower Upper
104 100
78 49.4 39.0 58.6
103 53.7 44.3 66.5









Bromoform
Concen: 53.07 ug/l
RT: 12.100 min Scan# 3222
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	173	Resp:	151565
Ion	Ratio	Lower	Upper
173	100		
175	48.1	24.3	72.8
254	0.1	0.0	0.0#

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM



1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. 0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion:	152	Resp:	223624
Ion	Ratio	Lower	Upper
152	100		
115	57.6	29.7	89.1
150	175.6	0.0	350.6



Abundance Scan 3260 (12.222 min): VN065092.D\data.ms (-3275) (-)

Isopropylbenzene
Concen: 51.45 ug/l
RT: 12.222 min Scan# 3260
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 794643
Ion Ratio Lower Upper
105 100
120 26.8 13.3 39.9


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 3204 (12.042 min): VN065092.D\data.ms (-3171) (-)


N-amyl acetate
Concen: 54.82 ug/l
RT: 12.042 min Scan# 3204
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion: 43 Resp: 288121
Ion Ratio Lower Upper
43 100
70 42.0 34.5 51.7
55 24.9 20.4 30.6
61 23.5 20.0 30.0



Abundance Scan 3339 (12.476 min): VN065092.D\data.ms (-3325) (-)





1,1,2,2-Tetrachloroethane
Concen: 50.07 ug/l
RT: 12.476 min Scan# 3339
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	83	Resp:	223077
Ion	Ratio	Lower	Upper
83	100		
131	10.6	5.1	15.3
85	64.0	33.0	99.0


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


MMDadoda
12/16/2020 1:41:11 PM



Abundance Scan 3339 (12.476 min): VN065113.D\data.ms (-3246) (-)





Abundance Scan 3355 (12.528 min): VN065092.D\data.ms (-3349) (-)





1,2,3-Trichloropropane
Concen: 52.99 ug/l m
RT: 12.527 min Scan# 3355
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	75	Resp:	212644
Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0



Abundance Scan 3355 (12.527 min): VN065113.D\data.ms (-3262) (-)









Bromobenzene
Concen: 51.37 ug/l
RT: 12.498 min Scan# 3346
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	156	Resp:	217051
Ion	Ratio	Lower	Upper
156	100		
77	179.6	101.3	303.7
158	96.3	48.2	144.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Manual Integrations
APPROVED


MMDadoda
12/16/2020 1:41:11 PM



n-propylbenzene
Concen: 52.02 ug/l
RT: 12.566 min Scan# 3367
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	91	Resp:	881058
Ion	Ratio	Lower	Upper
91	100		
120	24.7	11.6	34.7









2-Chlorotoluene
Concen: 50.83 ug/l
RT: 12.650 min Scan# 3393
Delta R.T. 0.004 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 91 Resp: 525363
Ion Ratio Lower Upper
91 100
126 35.7 17.2 51.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM



1,3,5-Trimethylbenzene
Concen: 53.42 ug/l
RT: 12.707 min Scan# 3411
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15


Tgt Ion:105 Resp: 680792
Ion Ratio Lower Upper
105 100
120 48.9 24.6 73.6



Abundance Scan 3276 (12.274 min): VN065092.D\data.ms (-3263) (-)

trans-1,4-Dichloro-2-butene
Concen: 51.68 ug/l
RT: 12.273 min Scan# 3276
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC



Tgt Ion:	75	Resp:	74528
Ion	Ratio	Lower	Upper
75	100		
53	94.1	67.0	100.4
89	48.2	39.6	59.4


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Abundance Scan 3423 (12.746 min): VN065092.D\data.ms (-3422) (-)

4-Chlorotoluene
Concen: 51.67 ug/l
RT: 12.746 min Scan# 3423
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15



Tgt Ion:	91	Resp:	552103
Ion	Ratio	Lower	Upper
91	100		
126	34.5	16.9	50.7









tert-Butylbenzene
Concen: 51.96 ug/l
RT: 12.968 min Scan# 3492
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	119	Resp:	570781
Ion	Ratio	Lower	Upper
119	100		
91	64.8	32.3	96.8
134	25.5	12.6	37.6

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC




Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM




1,2,4-Trimethylbenzene
Concen: 52.81 ug/l
RT: 13.013 min Scan# 3506
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	105	Resp:	679715
Ion	Ratio	Lower	Upper
105	100		
120	46.5	23.1	69.2



Abundance Scan 3547 (13.145 min): VN065092.D\data.ms (-3547) (-)






sec-Butylbenzene
Concen: 52.76 ug/l
RT: 13.145 min Scan# 3547
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:105 Resp: 734405
Ion Ratio Lower Upper
105 100
134 20.7 10.1 30.1


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM







Abundance Scan 3584 (13.264 min): VN065092.D\data.ms (-3584) (-)



p-Isopropyltoluene
Concen: 48.55 ug/l
RT: 13.264 min Scan# 3584
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:119 Resp: 690080
Ion Ratio Lower Upper
119 100
134 26.5 13.1 39.1
91 23.5 11.5 34.5









Scan 3581 (13.254 min): VN065092.D\data.ms (-3587) (-)
1,3-Dichlorobenzene
 Concen: 51.08 ug/l
 RT: 13.257 min Scan# 3582
 Delta R.T. 0.003 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion:	146	Resp:	372415
Ion	Ratio	Lower	Upper
146	100		
111	37.8	20.6	61.9
148	63.0	31.9	95.5

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC




Manual Integrations
 APPROVED

MMDadoda
 12/16/2020 1:41:11 PM




Scan 3606 (13.335 min): VN065092.D\data.ms (-3594) (-)
1,4-Dichlorobenzene
 Concen: 49.22 ug/l
 RT: 13.338 min Scan# 3607
 Delta R.T. 0.003 min
 Lab File: VN065113.D
 Acq: 14 Dec 2020 20:15

Tgt Ion:	146	Resp:	363070
Ion	Ratio	Lower	Upper
146	100		
111	38.4	19.8	59.3
148	63.6	31.9	95.5



Abundance Scan 3685 (13.589 min): VN065092.D\data.ms (-3679) (-)






n-Butylbenzene
Concen: 51.07 ug/l
RT: 13.588 min Scan# 3685
Delta R.T. -0.001 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion: 91 Resp: 552830
Ion Ratio Lower Upper
91 100
92 51.2 26.0 78.0
134 26.8 12.9 38.6


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 3765 (13.846 min): VN065092.D\data.ms (-3750) (-)




Hexachloroethane
Concen: 51.27 ug/l
RT: 13.846 min Scan# 3765
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:117 Resp: 120044
Ion Ratio Lower Upper
117 100
201 93.6 40.1 120.3



Abundance Scan 3697 (13.627 min): VN065092.D\data.ms (-3691) (-)






1,2-Dichlorobenzene
Concen: 49.58 ug/l
RT: 13.627 min Scan# 3697
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	146	Resp:	360384
Ion	Ratio	Lower	Upper
146	100		
111	39.4	21.1	63.1
148	63.2	32.0	96.2


Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM






Abundance Scan 3888 (14.241 min): VN065092.D\data.ms (-3891) (-)




1,2-Dibromo-3-Chloropropane
Concen: 50.59 ug/l
RT: 14.244 min Scan# 3889
Delta R.T. 0.003 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion:	75	Resp:	45447
Ion	Ratio	Lower	Upper
75	100		
155	87.1	42.8	128.3
157	112.0	56.4	169.1



Abundance Scan 4087 (14.881 min): VN065092.D\data.ms (-4093) (-)

1,2,4-Trichlorobenzene
Concen: 44.10 ug/l
RT: 14.881 min Scan# 4087
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15





Tgt Ion:	180	Resp:	191909
Ion Ratio	Lower	Upper	
180	100		
182	96.4	47.6	142.9
145	29.9	15.3	45.8

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM

Abundance Scan 4087 (14.881 min): VN065113.D\data.ms




Abundance Scan 4087 (14.881 min): VN065113.D\data.ms (-3994) (-)





Abundance Scan 4118 (14.981 min): VN065092.D\data.ms (-4104) (-)


Hexachlorobutadiene
Concen: 46.49 ug/l
RT: 14.981 min Scan# 4118
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15




Tgt Ion:	225	Resp:	114841
Ion Ratio	Lower	Upper	
225	100		
223	63.6	31.0	93.0
227	64.3	32.0	96.0



Abundance Scan 4118 (14.981 min): VN065113.D\data.ms (-4025) (-)









Naphthalene
Concen: 45.46 ug/l
RT: 15.103 min Scan# 4156
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
128	100		
127	12.7	10.3	15.5
129	10.8	8.7	13.1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

MMDadoda
12/16/2020 1:41:11 PM



1,2,3-Trichlorobenzene
Concen: 45.13 ug/l
RT: 15.283 min Scan# 4212
Delta R.T. -0.000 min
Lab File: VN065113.D
Acq: 14 Dec 2020 20:15

Tgt Ion	Ratio	Lower	Upper
180	100		
182	96.6	48.1	144.4
145	31.3	16.4	49.1

