

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070072.D
 Acq On : 14 Dec 2021 15:35
 Operator : JC/MD
 Sample : M5029-03
 Misc : 4.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PSC-124019

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/15/2021
 Supervised By : Mahesh Dadoda 12/21/2021

Quant Time: Dec 15 05:18:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.078	168	1070394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	2069191	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2403622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1125703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	875040	52.946	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.900%			
35) Dibromofluoromethane	8.016	113	562659	43.412	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 86.820%			
50) Toluene-d8	10.440	98	2827189	54.325	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.640%			
62) 4-Bromofluorobenzene	12.733	95	1214235	64.402	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 128.800%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.302	59	49095m	15.716	ug/l	
16) Acetone	4.307	43	17865	1.797	ug/l	99
25) 2-Butanone	7.348	43	146215	10.824	ug/l #	89
95) Naphthalene	15.507	128	101849	5.689	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070072.D
Acq On : 14 Dec 2021 15:35
Operator : JC/MD
Sample : M5029-03
Misc : 4.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PSC-124019

Quant Time: Dec 15 05:18:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/15/2021
Supervised By :Mahesh Dadoda 12/21/2021

