

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070077.D
 Acq On : 14 Dec 2021 17:58
 Operator : JC/MD
 Sample : M4990-12RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R8-6RE

Quant Time: Dec 15 05:20:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1186346	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2270222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2677575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1079843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	949384	51.830	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	103.660%	
35) Dibromofluoromethane	8.024	113	654267	46.010	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	92.020%	
50) Toluene-d8	10.443	98	3059603	53.584	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	107.160%	
62) 4-Bromofluorobenzene	12.731	95	1295157	62.611	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	125.220%	
Target Compounds						Qvalue
16) Acetone	4.299	43	81610	7.407	ug/l	98
20) Methylene Chloride	5.100	84	59644	4.072	ug/l	88
43) Isopropyl Acetate	8.566	43	137829	2.896	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070077.D
Acq On : 14 Dec 2021 17:58
Operator : JC/MD
Sample : M4990-12RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R8-6RE

Quant Time: Dec 15 05:20:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

