

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070078.D
 Acq On : 14 Dec 2021 18:23
 Operator : JC/MD
 Sample : M4990-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R3-7RE

Quant Time: Dec 15 05:20:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1035190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2021420	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2349593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	966905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	845937	52.926	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.860%
35) Dibromofluoromethane	8.024	113	583490	46.083	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.160%
50) Toluene-d8	10.443	98	2696374	53.035	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.080%
62) 4-Bromofluorobenzene	12.731	95	1136227	61.689	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.380%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	78355	8.149	ug/l	96
20) Methylene Chloride	5.093	84	57905	4.530	ug/l #	85
43) Isopropyl Acetate	8.563	43	144908	3.419	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070078.D
Acq On : 14 Dec 2021 18:23
Operator : JC/MD
Sample : M4990-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R3-7RE

Quant Time: Dec 15 05:20:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

