

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070079.D
 Acq On : 14 Dec 2021 18:49
 Operator : JC/MD
 Sample : M4989-11RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P28-10RE

Quant Time: Dec 15 05:21:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1165203	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2276143	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2671271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1131467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	949880	52.798	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.600%
35) Dibromofluoromethane	8.024	113	658144	46.162	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.320%
50) Toluene-d8	10.440	98	3032897	52.979	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.960%
62) 4-Bromofluorobenzene	12.731	95	1324890	63.882	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.760%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	82900	7.660	ug/l	95
20) Methylene Chloride	5.095	84	59337	4.124	ug/l	87
43) Isopropyl Acetate	8.563	43	159636	3.345	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070079.D
Acq On : 14 Dec 2021 18:49
Operator : JC/MD
Sample : M4989-11RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P28-10RE

Quant Time: Dec 15 05:21:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

