

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070080.D
 Acq On : 14 Dec 2021 19:15
 Operator : JC/MD
 Sample : M4989-04RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P23-2RE

Quant Time: Dec 15 05:21:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.085	168	1120957	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2194183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2580583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1087117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	905687	52.329	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.660%
35) Dibromofluoromethane	8.024	113	635021	46.204	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.400%
50) Toluene-d8	10.440	98	2914640	52.815	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.620%
62) 4-Bromofluorobenzene	12.731	95	1275937	63.820	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	127.640%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	76038	7.303	ug/l	90
20) Methylene Chloride	5.100	84	58122	4.199	ug/l #	81
43) Isopropyl Acetate	8.565	43	149688	3.254	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070080.D
Acq On : 14 Dec 2021 19:15
Operator : JC/MD
Sample : M4989-04RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P23-2RE

Quant Time: Dec 15 05:21:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

