

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070089.D
 Acq On : 14 Dec 2021 23:30
 Operator : JC/MD
 Sample : VN1214MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1214MBL02

Quant Time: Dec 15 05:25:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1077382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2058082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2373001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	954524	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	865834	52.050	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.100%
35) Dibromofluoromethane	8.024	113	595395	46.185	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.380%
50) Toluene-d8	10.440	98	2743355	52.998	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.000%
62) 4-Bromofluorobenzene	12.731	95	1162317	61.981	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070089.D
Acq On : 14 Dec 2021 23:30
Operator : JC/MD
Sample : VN1214MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1214MBL02

Quant Time: Dec 15 05:25:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

