

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070094.D
 Acq On : 15 Dec 2021 1:37
 Operator : JC/MD
 Sample : M5011-03 10X
 Misc : 3.72g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HSB-2-(9-9.5)

Quant Time: Dec 15 05:28:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	1215545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2330087	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2846383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1227875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	980675	52.252	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.500%			
35) Dibromofluoromethane	8.021	113	676273	46.335	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 92.680%			
50) Toluene-d8	10.440	98	3160111	53.923	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.840%			
62) 4-Bromofluorobenzene	12.731	95	1483624	69.880	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 139.760%#			
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.459	83	22730	0.861	ug/l #	67
67) Ethyl Benzene	11.843	91	210869	2.010	ug/l	96
68) m/p-Xylenes	11.950	106	199039	5.212	ug/l	92
69) o-Xylene	12.278	106	11459	0.298	ug/l	88
73) Isopropylbenzene	12.578	105	41267	0.378	ug/l	98
78) n-propylbenzene	12.921	91	208489	1.715	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	292606	3.328	ug/l	97
84) 1,2,4-Trimethylbenzene	13.366	105	905107	10.808	ug/l	97
85) sec-Butylbenzene	13.501	105	26223m	0.248	ug/l	
95) Naphthalene	15.504	128	91733	5.392	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070094.D
Acq On : 15 Dec 2021 1:37
Operator : JC/MD
Sample : M5011-03 10X
Misc : 3.72g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HSB-2-(9-9.5)

Quant Time: Dec 15 05:28:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

