

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070096.D
 Acq On : 15 Dec 2021 2:28
 Operator : JC/MD
 Sample : M5011-06 10X
 Misc : 2.37g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HSB-3-(11-11.5)

Quant Time: Dec 15 05:29:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1185071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2264604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2753236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1155737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	954014	52.139	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.280%	
35) Dibromofluoromethane	8.021	113	655897	46.239	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 92.480%	
50) Toluene-d8	10.440	98	3073285	53.958	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.920%	
62) 4-Bromofluorobenzene	12.731	95	1403017	67.994	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 135.980%#	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.318	59	24420	7.061	ug/l #	71
20) Methylene Chloride	5.098	84	5571	0.381	ug/l #	88
78) n-propylbenzene	12.921	91	34604	0.302	ug/l	96
80) 1,3,5-Trimethylbenzene	13.058	105	63742	0.770	ug/l	97
84) 1,2,4-Trimethylbenzene	13.366	105	212358	2.694	ug/l	95
95) Naphthalene	15.504	128	43870	4.700	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070096.D
Acq On : 15 Dec 2021 2:28
Operator : JC/MD
Sample : M5011-06 10X
Misc : 2.37g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HSB-3-(11-11.5)

Quant Time: Dec 15 05:29:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

