

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
 Data File : VN070098.D
 Acq On : 15 Dec 2021 3:19
 Operator : JC/MD
 Sample : M5011-10 10X
 Misc : 3.85g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HSB-5-(9-9.5)

Quant Time: Dec 15 05:30:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1192661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2304375	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2805546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1095796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	960925	52.183	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.360%			
35) Dibromofluoromethane	8.021	113	658466	45.619	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 91.240%			
50) Toluene-d8	10.440	98	3104318	53.562	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.120%			
62) 4-Bromofluorobenzene	12.731	95	1375096	65.491	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 130.980%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.321	59	37474	10.766	ug/l #	72
16) Acetone	4.293	43	12136	1.096	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121421\
Data File : VN070098.D
Acq On : 15 Dec 2021 3:19
Operator : JC/MD
Sample : M5011-10 10X
Misc : 3.85g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HSB-5-(9-9.5)

Quant Time: Dec 15 05:30:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

