

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
 Data File : VN075761.D
 Acq On : 14 Dec 2022 09:45
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 15 00:24:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	295335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	464371	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	388812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	149973	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	154929	48.220	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.440%
35) Dibromofluoromethane	8.171	113	141004	49.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	10.571	98	530745	49.361	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.720%
62) 4-Bromofluorobenzene	12.853	95	151747	42.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.220%
Target Compounds						
20) Methylene Chloride	5.283	84	1161	Below Cal	Qvalue #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
Data File : VN075761.D
Acq On : 14 Dec 2022 09:45
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 15 00:24:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

