

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
 Data File : VN075771.D
 Acq On : 14 Dec 2022 15:39
 Operator : JC\MD
 Sample : VN1214WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1214WBL01

Quant Time: Dec 15 00:27:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	343913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	528581	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	461064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	196045	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	175330	46.861	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.720%
35) Dibromofluoromethane	8.177	113	159629	48.765	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	10.571	98	614098	50.176	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.360%
62) 4-Bromofluorobenzene	12.853	95	188909	46.602	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
Data File : VN075771.D
Acq On : 14 Dec 2022 15:39
Operator : JC\MD
Sample : VN1214WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1214WBL01

Quant Time: Dec 15 00:27:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

