

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
 Data File : VN075773.D
 Acq On : 14 Dec 2022 16:27
 Operator : JC\MD
 Sample : N5985-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-02-120922

Quant Time: Dec 15 00:27:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	323900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	497409	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	402816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	165075	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159588	45.290	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.580%
35) Dibromofluoromethane	8.171	113	144051	46.764	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.520%
50) Toluene-d8	10.571	98	562650	48.853	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.700%
62) 4-Bromofluorobenzene	12.853	95	158934	41.664	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
Data File : VN075773.D
Acq On : 14 Dec 2022 16:27
Operator : JC\MD
Sample : N5985-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-02-120922

Quant Time: Dec 15 00:27:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

