

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
 Data File : VN075776.D
 Acq On : 14 Dec 2022 17:39
 Operator : JC\MD
 Sample : N6001-07DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7DL

Quant Time: Dec 15 00:28:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	324589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	505410	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	442728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191906	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	164716	46.646	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.300%
35) Dibromofluoromethane	8.171	113	149273	47.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	10.571	98	580760	49.627	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.260%
62) 4-Bromofluorobenzene	12.853	95	186847	48.206	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.420%
Target Compounds						
					Qvalue	
40) Benzene	8.612	78	129563	10.100	ug/l	99
52) Toluene	10.635	92	40309	4.825	ug/l	98
67) Ethyl Benzene	11.965	91	424460	27.677	ug/l	98
68) m/p-Xylenes	12.070	106	350555	57.751	ug/l	100
69) o-Xylene	12.400	106	66831	11.125	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
Data File : VN075776.D
Acq On : 14 Dec 2022 17:39
Operator : JC\MD
Sample : N6001-07DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7DL

Quant Time: Dec 15 00:28:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

