

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
 Data File : VN075778.D
 Acq On : 14 Dec 2022 18:27
 Operator : JC\MD
 Sample : N6026-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-12-12

Quant Time: Dec 15 00:29:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	353864	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	545503	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	474698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	200720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	177288	46.052	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.100%
35) Dibromofluoromethane	8.171	113	162645	48.145	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.300%
50) Toluene-d8	10.571	98	631083	49.964	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.920%
62) 4-Bromofluorobenzene	12.853	95	197220	47.143	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.280%
Target Compounds						
68) m/p-Xylenes	12.071	106	2600	0.399	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
Data File : VN075778.D
Acq On : 14 Dec 2022 18:27
Operator : JC\MD
Sample : N6026-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-12-12

Quant Time: Dec 15 00:29:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

