

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
 Data File : VN075779.D
 Acq On : 14 Dec 2022 18:51
 Operator : JC\MD
 Sample : N6001-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6M

Quant Time: Dec 15 00:29:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	360374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	559644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	479958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	207699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	181397	46.268	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.540%		
35) Dibromofluoromethane	8.171	113	165405	47.725	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.440%		
50) Toluene-d8	10.571	98	638261	49.255	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.520%		
62) 4-Bromofluorobenzene	12.853	95	197578	46.035	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.080%		
Target Compounds						Qvalue
30) Chloroform	7.971	83	8658	1.271	ug/l	91
40) Benzene	8.612	78	15798	1.112	ug/l #	89
68) m/p-Xylenes	12.076	106	2339	0.355	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121422\
Data File : VN075779.D
Acq On : 14 Dec 2022 18:51
Operator : JC\MD
Sample : N6001-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6M

Quant Time: Dec 15 00:29:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

