

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
 Data File : VN080479.D
 Acq On : 14 Dec 2023 12:35
 Operator : JC\MD
 Sample : VN1214WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1214WBL01

Quant Time: Dec 15 00:25:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	238949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	429372	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	440079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201622	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	157600	51.640	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.280%
35) Dibromofluoromethane	8.165	113	129867	47.703	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.400%
50) Toluene-d8	10.565	98	472177	46.506	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.020%
62) 4-Bromofluorobenzene	12.847	95	189571	49.170	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
Data File : VN080479.D
Acq On : 14 Dec 2023 12:35
Operator : JC\MD
Sample : VN1214WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1214WBL01

Quant Time: Dec 15 00:25:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

