

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
 Data File : VN080489.D
 Acq On : 14 Dec 2023 17:50
 Operator : JC\MD
 Sample : 05791-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 597-P-2

Quant Time: Dec 15 00:31:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	243799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	442266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	463595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	211469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	168653	54.162	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	8.165	113	138975	49.561	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.120%	
50) Toluene-d8	10.565	98	508873	48.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.320%	
62) 4-Bromofluorobenzene	12.853	95	209403	52.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.460%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	57219	103.634	ug/l	97
20) Methylene Chloride	5.277	84	18830	6.373	ug/l	87
25) 2-Butanone	7.482	43	5742	5.399	ug/l	92
37) Ethyl Acetate	7.559	43	2913	1.015	ug/l #	86
43) Isopropyl Acetate	8.688	43	123089	23.844	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
Data File : VN080489.D
Acq On : 14 Dec 2023 17:50
Operator : JC\MD
Sample : 05791-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
597-P-2

Quant Time: Dec 15 00:31:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

