

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
 Data File : VN080490.D
 Acq On : 14 Dec 2023 18:13
 Operator : JC\MD
 Sample : 05791-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1B-SOIL

Quant Time: Dec 15 00:31:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	239689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	445507	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	474405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	217273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	168132	54.920	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.840%	
35) Dibromofluoromethane	8.165	113	140051	49.581	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.160%	
50) Toluene-d8	10.570	98	505856	48.018	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.040%	
62) 4-Bromofluorobenzene	12.847	95	214163	53.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.080%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	58044	106.931	ug/l	95
20) Methylene Chloride	5.283	84	19588	6.743	ug/l	94
25) 2-Butanone	7.494	43	5958	5.698	ug/l	99
43) Isopropyl Acetate	8.688	43	121852	23.432	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
Data File : VN080490.D
Acq On : 14 Dec 2023 18:13
Operator : JC\MD
Sample : 05791-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1B-SOIL

Quant Time: Dec 15 00:31:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

