

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
 Data File : VN080491.D
 Acq On : 14 Dec 2023 18:37
 Operator : JC\MD
 Sample : 05803-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 600572408-

Quant Time: Dec 15 00:32:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	233087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	434545	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	448956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	206787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	162411	54.554	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.100%	
35) Dibromofluoromethane	8.171	113	133462	48.440	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.880%	
50) Toluene-d8	10.565	98	491017	47.786	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.580%	
62) 4-Bromofluorobenzene	12.847	95	199815	51.210	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.420%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	52906	100.226	ug/l	98
20) Methylene Chloride	5.271	84	18553	6.568	ug/l	96
25) 2-Butanone	7.483	43	5708	5.613	ug/l	88
43) Isopropyl Acetate	8.688	43	133250	26.271	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121423\
Data File : VN080491.D
Acq On : 14 Dec 2023 18:37
Operator : JC\MD
Sample : 05803-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
600572408-

Quant Time: Dec 15 00:32:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

