

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
 Data File : VN070105.D
 Acq On : 15 Dec 2021 11:50
 Operator : JC/MD
 Sample : VN1215MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1215MBL01

Quant Time: Dec 16 05:37:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1185126	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2280571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2690784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1152076	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	957287	52.315	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.640%
35) Dibromofluoromethane	8.021	113	668921	46.827	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.660%
50) Toluene-d8	10.443	98	3030077	52.827	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.660%
62) 4-Bromofluorobenzene	12.731	95	1357356	65.321	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	130.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
Data File : VN070105.D
Acq On : 15 Dec 2021 11:50
Operator : JC/MD
Sample : VN1215MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1215MBL01

Quant Time: Dec 16 05:37:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

