

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
 Data File : VN070110.D
 Acq On : 15 Dec 2021 14:36
 Operator : JC/MD
 Sample : M4970-03DL 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-21DL

Quant Time: Dec 16 05:38:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1046972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1992304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2290844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	935595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	835397	51.679	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.360%
35) Dibromofluoromethane	8.021	113	573988	45.995	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.980%
50) Toluene-d8	10.441	98	2631509	52.516	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.040%
62) 4-Bromofluorobenzene	12.731	95	1130744	62.289	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	124.580%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.380	59	171911	56.260	ug/l	99
40) Benzene	8.464	78	12162	0.210	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
Data File : VN070110.D
Acq On : 15 Dec 2021 14:36
Operator : JC/MD
Sample : M4970-03DL 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-21DL

Quant Time: Dec 16 05:38:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

