

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
 Data File : VN070114.D
 Acq On : 15 Dec 2021 16:18
 Operator : JC/MD
 Sample : M4970-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-1

Quant Time: Dec 16 05:39:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1168251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2225097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2609013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1102649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	942742	52.265	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.520%	
35) Dibromofluoromethane	8.024	113	648595	46.536	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 93.080%	
50) Toluene-d8	10.443	98	2954470	52.793	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.580%	
62) 4-Bromofluorobenzene	12.733	95	1325084	65.357	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 130.720%	
Target Compounds						Qvalue
11) Tert butyl alcohol	5.379	59	350412	102.773	ug/l	99
16) Acetone	4.291	43	479394	44.182	ug/l	96
19) Methyl tert-butyl Ether	5.621	73	173100	4.001	ug/l	92
22) Diisopropyl ether	6.500	45	170144	3.471	ug/l #	87
25) 2-Butanone	7.343	43	68434	4.642	ug/l #	89
40) Benzene	8.461	78	1480262	22.863	ug/l	99
52) Toluene	10.507	92	16326	0.413	ug/l	91
78) n-propylbenzene	12.921	91	27113	0.248	ug/l	93
86) p-Isopropyltoluene	13.616	119	87342	1.178	ug/l	94
95) Naphthalene	15.507	128	50916	4.855	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
Data File : VN070114.D
Acq On : 15 Dec 2021 16:18
Operator : JC/MD
Sample : M4970-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-1

Quant Time: Dec 16 05:39:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

