

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
 Data File : VN070124.D
 Acq On : 15 Dec 2021 20:34
 Operator : JC/MD
 Sample : M5078-04
 Misc : 4.07g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RB-21141

Quant Time: Dec 16 05:41:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1149969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2203180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2594215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1198060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	914458	51.503	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	103.000%	
35) Dibromofluoromethane	8.018	113	602185	43.636	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	87.280%	
50) Toluene-d8	10.440	98	2998070	54.105	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	108.200%	
62) 4-Bromofluorobenzene	12.733	95	1335248	66.514	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	133.020%	
Target Compounds						
						Qvalue
16) Acetone	4.304	43	13652	1.278	ug/l	97
18) Methyl Acetate	4.870	43	16865	0.523	ug/l #	80
25) 2-Butanone	7.353	43	14219	0.980	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121521\
Data File : VN070124.D
Acq On : 15 Dec 2021 20:34
Operator : JC/MD
Sample : M5078-04
Misc : 4.07g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RB-21141

Quant Time: Dec 16 05:41:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

