

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
 Data File : VN075813.D
 Acq On : 15 Dec 2022 13:24
 Operator : JC\MD
 Sample : N5990-11DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP120722-2DL

Quant Time: Dec 16 03:08:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	358261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	557605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	480568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	196828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	183539	47.091	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.180%
35) Dibromofluoromethane	8.177	113	163480	47.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.680%
50) Toluene-d8	10.571	98	637036	49.341	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.853	95	194867	45.569	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.140%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.530	62	8705	2.307	ug/l	95
27) cis-1,2-Dichloroethene	7.500	96	31520	7.579	ug/l	90
44) Trichloroethene	9.353	130	11825	2.966	ug/l	100
64) Tetrachloroethene	11.106	164	68508	20.259	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
Data File : VN075813.D
Acq On : 15 Dec 2022 13:24
Operator : JC\MD
Sample : N5990-11DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP120722-2DL

Quant Time: Dec 16 03:08:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

