

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
 Data File : VN075820.D
 Acq On : 15 Dec 2022 16:11
 Operator : JC\MD
 Sample : N6074-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 221214021-01-TRIP-BLANK

Quant Time: Dec 16 03:10:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	354543	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	550993	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	478918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	182504	47.316	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.640%		
35) Dibromofluoromethane	8.177	113	167574	49.110	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.220%		
50) Toluene-d8	10.571	98	634481	49.732	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.460%		
62) 4-Bromofluorobenzene	12.853	95	197437	46.725	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.440%		
Target Compounds					Qvalue	
16) Acetone	4.459	43	3896	2.930	ug/l	97
30) Chloroform	7.971	83	3850	0.575	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
Data File : VN075820.D
Acq On : 15 Dec 2022 16:11
Operator : JC\MD
Sample : N6074-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
221214021-01-TRIP-BLANK

Quant Time: Dec 16 03:10:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

