

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
 Data File : VN075823.D
 Acq On : 15 Dec 2022 17:23
 Operator : JC\MD
 Sample : N6067-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-LIQUID-20221214

Quant Time: Dec 16 03:11:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	339985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	533522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	466489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	201864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	177122	47.887	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.780%
35) Dibromofluoromethane	8.177	113	161069	48.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.500%
50) Toluene-d8	10.570	98	616794	49.929	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.860%
62) 4-Bromofluorobenzene	12.853	95	197670	48.312	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.620%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	21908	17.182	ug/l	91
25) 2-Butanone	7.494	43	5459	3.058	ug/l #	73
40) Benzene	8.612	78	6853	0.506	ug/l #	86
46) Dibromomethane	9.706	93	804	0.329	ug/l #	85
68) m/p-Xylenes	12.070	106	4177	0.653	ug/l	90
69) o-Xylene	12.394	106	2321	0.367	ug/l	82
84) 1,2,4-Trimethylbenzene	13.482	105	3632	0.300	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
Data File : VN075823.D
Acq On : 15 Dec 2022 17:23
Operator : JC\MD
Sample : N6067-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-LIQUID-20221214

Quant Time: Dec 16 03:11:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

