

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
 Data File : VN075825.D
 Acq On : 15 Dec 2022 18:11
 Operator : JC\MD
 Sample : N6066-06 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ORA-594

Quant Time: Dec 16 03:12:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	373121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	575991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	511247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	223457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	192414	47.402	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.800%
35) Dibromofluoromethane	8.177	113	174819	49.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.020%
50) Toluene-d8	10.570	98	671088	50.319	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.640%
62) 4-Bromofluorobenzene	12.853	95	214245	48.502	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%
Target Compounds						
16) Acetone	4.453	43	32805	23.444	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
Data File : VN075825.D
Acq On : 15 Dec 2022 18:11
Operator : JC\MD
Sample : N6066-06 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ORA-594

Quant Time: Dec 16 03:12:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

