

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
 Data File : VN075826.D
 Acq On : 15 Dec 2022 18:35
 Operator : JC\MD
 Sample : N6066-07 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ORA-595

Quant Time: Dec 16 03:12:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	363419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	560186	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	481122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	192516	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	187470	47.417	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.840%
35) Dibromofluoromethane	8.171	113	168067	48.446	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	10.571	98	651066	50.195	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.380%
62) 4-Bromofluorobenzene	12.853	95	197247	45.913	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121522\
Data File : VN075826.D
Acq On : 15 Dec 2022 18:35
Operator : JC\MD
Sample : N6066-07 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ORA-595

Quant Time: Dec 16 03:12:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

