

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121523\
 Data File : VN080496.D
 Acq On : 15 Dec 2023 10:27
 Operator : JC\MD
 Sample : VN1215MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 VN1215MBL01

Quant Time: Dec 16 02:49:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	8.224	168	226693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	407169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	418088	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	189274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142731	49.296	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.600%
35) Dibromofluoromethane	8.159	113	121172	46.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.880%
50) Toluene-d8	10.565	98	418917	43.510	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.020%
62) 4-Bromofluorobenzene	12.847	95	170458	46.624	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121523\
 Data File : VN080496.D
 Acq On : 15 Dec 2023 10:27
 Operator : JC\MD
 Sample : VN1215MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1215MBL01

Quant Time: Dec 16 02:49:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

