

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121523\
 Data File : VN080502.D
 Acq On : 15 Dec 2023 14:48
 Operator : JC\MD
 Sample : PB157817TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157817TB

Quant Time: Dec 16 02:52:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	232183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	425082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	445273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	203673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	158587	53.477	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.960%	
35) Dibromofluoromethane	8.165	113	135905	50.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.860%	
50) Toluene-d8	10.570	98	485779	48.328	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.660%	
62) 4-Bromofluorobenzene	12.847	95	203635	53.351	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.700%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.518	59	4527	12.962	ug/l #	76
16) Acetone	4.424	43	30028	57.107	ug/l	94
20) Methylene Chloride	5.277	84	19741	7.015	ug/l	91
25) 2-Butanone	7.482	43	17673	17.447	ug/l	89
43) Isopropyl Acetate	8.688	43	122110	24.610	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121523\
Data File : VN080502.D
Acq On : 15 Dec 2023 14:48
Operator : JC\MD
Sample : PB157817TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157817TB

Quant Time: Dec 16 02:52:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

