

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121523\
 Data File : VN080506.D
 Acq On : 15 Dec 2023 16:23
 Operator : JC\MD
 Sample : 05862-12DL 10X
 Misc : 5.1mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-LIQUIDS-20231213-01DL

Quant Time: Dec 16 02:55:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 13 01:57:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	238867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	435383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	463867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	235218	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	162856	53.380	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.760%
35) Dibromofluoromethane	8.165	113	139381	50.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.980%
50) Toluene-d8	10.565	98	517042	50.221	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.440%
62) 4-Bromofluorobenzene	12.853	95	225415	57.660	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.320%

Target Compounds					Qvalue	
16) Acetone	4.442	43	2691	4.975	ug/l	# 80
40) Benzene	8.612	78	15364	1.179	ug/l	91
52) Toluene	10.635	92	23221	2.820	ug/l	96
67) Ethyl Benzene	11.965	91	689873	36.359	ug/l	97
68) m/p-Xylenes	12.071	106	222377	31.717	ug/l	92
69) o-Xylene	12.400	106	121223	17.677	ug/l	88
73) Isopropylbenzene	12.694	105	34538	1.828	ug/l	98
80) 1,3,5-Trimethylbenzene	13.171	105	35397	2.175	ug/l	95
84) 1,2,4-Trimethylbenzene	13.482	105	156274	9.653	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121523\
Data File : VN080506.D
Acq On : 15 Dec 2023 16:23
Operator : JC\MD
Sample : 05862-12DL 10X
Misc : 5.1mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-LIQUIDS-20231213-01DL

Quant Time: Dec 16 02:55:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121223W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 13 01:57:14 2023
Response via : Initial Calibration

