

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121619\
 Data File : VN059610.D
 Acq On : 16 Dec 2019 18:07
 Operator : JC/SP
 Sample : PB125470TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB125470TB

Quant Time: Dec 17 04:53:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	228525	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	370104	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	331765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	130598	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	166174	49.97	ug/l	0.00
Spiked Amount			Recovery	=	99.94%	
35) Dibromofluoromethane	7.57	113	115421	49.84	ug/l	0.00
Spiked Amount			Recovery	=	99.68%	
50) Toluene-d8	10.08	98	465679	50.79	ug/l	0.00
Spiked Amount			Recovery	=	101.58%	
62) 4-Bromofluorobenzene	12.40	95	152928	45.71	ug/l	0.00
Spiked Amount			Recovery	=	91.42%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.75	59	715	1.173	ug/l #	72
16) Acetone	3.79	43	12646	6.405	ug/l	96
17) Carbon Disulfide	4.02	76	1487	2.051	ug/l #	75
18) Methyl Acetate	4.30	43	4024	0.973	ug/l #	78
20) Methylene Chloride	4.52	84	2649	0.929	ug/l #	93
25) 2-Butanone	6.82	43	6170	2.613	ug/l #	89
29) Tetrahydrofuran	7.19	42	473	0.329	ug/l #	35
31) Cyclohexane	7.64	56	5166	1.064	ug/l #	8
36) 1,1-Dichloropropene	7.64	75	18905	4.898	ug/l #	49
38) Carbon Tetrachloride	7.64	117	23510	5.870	ug/l #	17
43) Isopropyl Acetate	8.15	43	71659	10.275	ug/l #	92
45) 1,2-Dichloropropane	9.25	63	1503	0.493	ug/l #	44
48) Methyl methacrylate	9.16	41	4518	1.428	ug/l #	13
51) 4-Methyl-2-Pentanone	9.98	43	5077	1.221	ug/l	93
52) Toluene	10.14	92	1473	0.213	ug/l #	72
54) cis-1,3-Dichloropropene	9.89	75	28919	5.979	ug/l #	63
56) Ethyl methacrylate	10.54	69	1134	0.281	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	4861	1.027	ug/l #	42
59) 2-Hexanone	10.76	43	82022	25.027	ug/l #	53
74) N-amyl acetate	12.03	43	4107	0.873	ug/l #	48
76) 1,2,3-Trichloropropane	12.40	75	82549	25.533	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.73	105	4028	0.464	ug/l	92
81) trans-1,4-Dichloro-2-buten	12.40	75	82549	63.064	ug/l #	11
92) 1,2-Dibromo-3-Chloropropan	14.26	75	177	1.813	ug/l #	6


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121619\
Data File : VN059610.D
Acq On : 16 Dec 2019 18:07
Operator : JC/SP
Sample : PB125470TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB125470TB

Quant Time: Dec 17 04:53:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.64 min Scan# 1825
Delta R.T. -0.00 min
Lab File: VN059610.D
Acq: 16 Dec 2019 18:07

Instrument :
MSVOA_N
ClientSampleId :
PB125470TB

Tgt Ion:168 Resp: 228525
Ion Ratio Lower Upper
168 100
99 60.1 48.2 72.4




Abundance Ion 167.90 (167.60 to 168.60): V
Ion 98.90 (98.60 to 99.60): VN






#11
Tert butyl alcohol
Concen: 1.173 ug/l
RT: 4.75 min Scan# 927
Delta R.T. 0.00 min
Lab File: VN059610.D
Acq: 16 Dec 2019 18:07

Tgt Ion: 59 Resp: 715
Ion Ratio Lower Upper
59 100
57 0.0 8.6 12.8#



Abundance Ion 59.00 (58.70 to 59.70): VN
Ion 57.00 (56.70 to 57.70): VN





#16

Acetone

Concen: 6.405 ug/l

RT: 3.79 min Scan# 627

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :


PB125470TB

Tgt Ion: 43 Resp: 12646

Ion Ratio Lower Upper




43 100

58 27.1 23.6 35.4



Abundance Ion 43.00 (42.70 to 43.70): VN059587.D (-593) (-)

Ion 58.00 (57.70 to 58.70): VN059587.D (-593) (-)



#17

Carbon Disulfide

Concen: 2.051 ug/l

RT: 4.02 min Scan# 699

Delta R.T. -0.00 min

Lab File: VN059610.D


Acq: 16 Dec 2019 18:07

Tgt Ion: 76 Resp: 1487

Ion Ratio Lower Upper



76 100


78 0.0 7.4 11.0#



Abundance Ion 75.90 (75.60 to 76.60): VN059587.D (-679) (-)

Ion 77.90 (77.60 to 78.60): VN059587.D (-679) (-)





#18

Methyl Acetate

Concen: 0.973 ug/l

RT: 4.30 min Scan# 785

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampled :


PB125470TB

Tgt Ion: 43 Resp: 4024

Ion Ratio Lower Upper




43 100

74 11.3 17.5 26.3#



Abundance Ion 43.00 (42.70 to 43.70): VN059610.D (-691) (-)

Ion 74.00 (73.70 to 74.70): VN059610.D (-691) (-)



#20

Methylene Chloride

Concen: 0.929 ug/l

RT: 4.52 min Scan# 854

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Tgt Ion: 84 Resp: 2649


Ion Ratio Lower Upper

84 100

49 136.3 106.6 160.0

51 40.6 31.9 47.9

86 79.7 50.8 76.2#






Abundance Ion 83.90 (83.60 to 84.60): VN059610.D (-761) (-)

Ion 48.90 (48.60 to 49.60): VN059610.D (-761) (-)

Ion 50.90 (50.60 to 51.60): VN059610.D (-761) (-)

Ion 85.90 (85.60 to 86.60): VN059610.D (-761) (-)





#25

2-Butanone

Concen: 2.613 ug/l

RT: 6.82 min Scan# 1568

Delta R.T. 0.01 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :





PB125470TB

Tgt Ion: 43 Resp: 6170

Ion Ratio Lower Upper

43 100

72 19.6 19.9 29.9#



#29

Tetrahydrofuran

Concen: 0.329 ug/l

RT: 7.19 min Scan# 1685

Delta R.T. 0.01 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07




Tgt Ion: 42 Resp: 473


Ion Ratio Lower Upper

42 100

72 0.0 33.1 49.7#

71 0.0 31.4 47.0#









#31
Cyclohexane
Concen: 1.064 ug/l
RT: 7.64 min Scan# 1824
Delta R.T. 0.01 min
Lab File: VN059610.D
Acq: 16 Dec 2019 18:07

Instrument :
MSVOA_N
ClientSampleId :
PB125470TB

Tgt Ion: 56 Resp: 5166
Ion Ratio Lower Upper
56 100
69 161.6 26.5 39.7#
84 114.5 65.3 97.9#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 49.971 ug/l
RT: 8.00 min Scan# 1938
Delta R.T. 0.00 min
Lab File: VN059610.D
Acq: 16 Dec 2019 18:07

Tgt Ion: 65 Resp: 166174
Ion Ratio Lower Upper
65 100
67 51.2 0.0 103.0



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2113
 Delta R.T. 0.00 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07

Instrument :
 MSVOA_N
 ClientSampleId :
 PB125470TB

Tgt Ion:	114	Resp:	370104
Ion	Ratio	Lower	Upper
114	100		
63	22.3	0.0	43.6
88	16.6	0.0	31.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 49.843 ug/l
 RT: 7.57 min Scan# 1802
 Delta R.T. 0.00 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07

Tgt Ion:	113	Resp:	115421
Ion	Ratio	Lower	Upper
113	100		
111	104.1	83.5	125.3
192	18.5	15.4	23.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.898 ug/l
 RT: 7.64 min Scan# 1825
 Delta R.T. -0.13 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07




Instrument :
 MSVOA_N
 ClientSampleId :
 PB125470TB


Tgt Ion: 75 Resp: 18905
 Ion Ratio Lower Upper
 75 100
 110 7.5 16.6 49.7#
 77 0.0 25.1 37.7#



#38
 Carbon Tetrachloride
 Concen: 5.870 ug/l
 RT: 7.64 min Scan# 1825
 Delta R.T. -0.11 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07

Tgt Ion: 117 Resp: 23510
 Ion Ratio Lower Upper
 117 100
 119 5.7 76.5 114.7#
 121 0.0 25.1 37.7#





#43

Isopropyl Acetate

Concen: 10.275 ug/l

RT: 8.15 min Scan# 1982

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :

PB125470TB





Tgt Ion: 43 Resp: 71659

Ion Ratio Lower Upper

43 100

61 19.2 20.2 30.2#

87 12.2 9.5 14.3



#45

1,2-Dichloropropane

Concen: 0.493 ug/l

RT: 9.25 min Scan# 2326

Delta R.T. 0.15 min

Lab File: VN059610.D




Acq: 16 Dec 2019 18:07


Tgt Ion: 63 Resp: 1503

Ion Ratio Lower Upper

63 100

65 0.0 24.9 37.3#









#48
Methyl methacrylate
Concen: 1.428 ug/l
RT: 9.16 min Scan# 2298
Delta R.T. -0.02 min
Lab File: VN059610.D
Acq: 16 Dec 2019 18:07




Instrument :
MSVOA_N
ClientSampleId :
PB125470TB


Tgt Ion: 41 Resp: 4518
Ion Ratio Lower Upper
41 100
69 0.0 64.4 96.6#
39 0.0 51.3 76.9#



#50
Toluene-d8
Concen: 50.793 ug/l
RT: 10.08 min Scan# 2583
Delta R.T. 0.00 min
Lab File: VN059610.D
Acq: 16 Dec 2019 18:07

Tgt Ion: 98 Resp: 465679
Ion Ratio Lower Upper
98 100
100 64.1 51.2 76.8









#51
 4-Methyl-2-Pentanone
 Concen: 1.221 ug/l
 RT: 9.98 min Scan# 2551
 Delta R.T. 0.00 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07




Instrument :
 MSVOA_N
 ClientSampled :
 PB125470TB


Tgt Ion: 43 Resp: 5077
 Ion Ratio Lower Upper
 43 100
 58 41.4 29.8 44.8



#52
 Toluene
 Concen: 0.213 ug/l
 RT: 10.14 min Scan# 2603
 Delta R.T. -0.00 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07

Tgt Ion: 92 Resp: 1473
 Ion Ratio Lower Upper
 92 100
 91 215.0 140.4 210.6#





#54

cis-1,3-Dichloropropene

Concen: 5.979 ug/l

RT: 9.89 min Scan# 2525

Delta R.T. 0.06 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :

PB125470TB





Tgt Ion: 75 Resp: 28919

Ion Ratio Lower Upper

75 100

77 0.0 25.1 37.7#

39 38.2 46.1 69.1#



#56

Ethyl methacrylate

Concen: 0.281 ug/l

RT: 10.54 min Scan# 2727

Delta R.T. 0.12 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07




Tgt Ion: 69 Resp: 1134


Ion Ratio Lower Upper

69 100

41 5611.4 57.1 85.7#

39 2510.6 34.7 52.1#









#57
 1,3-Dichloropropane
 Concen: 1.027 ug/l
 RT: 10.76 min Scan# 2794
 Delta R.T. 0.06 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07




Instrument :
 MSVOA_N
 ClientSampleId :
 PB125470TB


Tgt Ion: 76 Resp: 4861
 Ion Ratio Lower Upper
 76 100
 78 0.0 25.8 38.8#



#59
 2-Hexanone
 Concen: 25.027 ug/l
 RT: 10.76 min Scan# 2794
 Delta R.T. 0.01 min
 Lab File: VN059610.D
 Acq: 16 Dec 2019 18:07

Tgt Ion: 43 Resp: 82022
 Ion Ratio Lower Upper
 43 100
 58 18.9 25.8 77.3#





#62

4-Bromofluorobenzene

Concen: 45.713 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :

PB125470TB


Tgt Ion: 95 Resp: 152928

Ion Ratio Lower Upper

95 100

174 75.8 0.0 151.4




176 74.5 0.0 146.0



Abundance Ion 94.90 (94.60 to 95.60): VN059610.D (-3212) (-)

Ion 173.80 (173.50 to 174.50): VN059610.D (-3212) (-)

Ion 175.80 (175.50 to 176.50): VN059610.D (-3212) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.40 min Scan# 2995

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07


Tgt Ion: 117 Resp: 331765

Ion Ratio Lower Upper

117 100

82 58.2 45.9 68.9



119 31.9 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): VN059610.D (-2902) (-)

Ion 82.00 (81.70 to 82.70): VN059610.D (-2902) (-)

Ion 118.90 (118.60 to 119.60): VN059610.D (-2902) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :

PB125470TB


Tgt Ion:152 Resp: 130598

Ion Ratio Lower Upper

152 100

115 60.2 30.3 90.9




150 157.9 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.873 ug/l

RT: 12.03 min Scan# 3189

Delta R.T. -0.04 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Tgt Ion: 43 Resp: 4107


Ion Ratio Lower Upper

43 100

70 0.0 31.3 46.9#

55 5.3 19.8 29.6#

61 0.0 18.3 27.5#






Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 70.00 (69.70 to 70.70): VNO

Ion 55.00 (54.70 to 55.70): VNO

Ion 61.00 (60.70 to 61.70): VNO





#76

1,2,3-Trichloropropane

Concen: 25.533 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampled :


PB125470TB

Tgt Ion: 75 Resp: 82549

Ion Ratio Lower Upper




75 100

77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN059610.D (-3259) (-)

Ion 77.00 (76.70 to 77.70): VN059610.D (-3259) (-)



#80

1,3,5-Trimethylbenzene

Concen: 0.464 ug/l

RT: 12.73 min Scan# 3409

Delta R.T. 0.00 min

Lab File: VN059610.D


Acq: 16 Dec 2019 18:07

Tgt Ion: 105 Resp: 4028

Ion Ratio Lower Upper



105 100


120 42.9 24.1 72.4



Abundance Ion 105.00 (104.70 to 105.70): VN059610.D (-3315) (-)

Ion 120.00 (119.70 to 120.70): VN059610.D (-3315) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 63.064 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Instrument :

MSVOA_N

ClientSampleId :

PB125470TB





Tgt Ion: 75 Resp: 82549

Ion Ratio Lower Upper

75 100

53 0.0 78.1 117.1#

89 0.0 35.4 53.2#



#92

1,2-Dibromo-3-Chloropropane

Concen: 1.813 ug/l

RT: 14.26 min Scan# 3885

Delta R.T. -0.01 min

Lab File: VN059610.D

Acq: 16 Dec 2019 18:07

Tgt Ion: 75 Resp: 177

Ion Ratio Lower Upper

75 100

155 0.0 37.0 110.9#

157 0.0 48.4 145.2#

