

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121619\
 Data File : VN059607.D
 Acq On : 16 Dec 2019 17:01
 Operator : JC/SP
 Sample : K6269-06DL 500X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 23810DL

Quant Time: Dec 17 06:10:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	251642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	402059	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	363624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	145749	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	180806	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	7.57	113	125562	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
50) Toluene-d8	10.08	98	498273	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	12.40	95	166203	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
Target Compounds						
16) Acetone	3.79	43	286239	131.663	ug/l	100
52) Toluene	10.15	92	11739	1.559	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121619\
Data File : VN059607.D
Acq On : 16 Dec 2019 17:01
Operator : JC/SP
Sample : K6269-06DL 500X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
23810DL

Quant Time: Dec 17 06:10:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

