

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121619\
 Data File : VN059612.D
 Acq On : 16 Dec 2019 18:51
 Operator : JC/SP
 Sample : K6112-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-03-121219

Quant Time: Dec 17 06:27:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 13 23:17:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	237134	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	381926	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	346156	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	138043	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	170445	49.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.78%	
35) Dibromofluoromethane	7.57	113	120251	50.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.64%	
50) Toluene-d8	10.08	98	478400	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
62) 4-Bromofluorobenzene	12.40	95	160269	46.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.84%	
Target Compounds						
16) Acetone	3.79	43	18674	9.115	ug/l	93
20) Methylene Chloride	4.52	84	288353	97.457	ug/l	99
30) Chloroform	7.35	83	6567	1.251	ug/l	95
43) Isopropyl Acetate	8.14	43	65299	9.073	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121619\
Data File : VN059612.D
Acq On : 16 Dec 2019 18:51
Operator : JC/SP
Sample : K6112-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-03-121219

Quant Time: Dec 17 06:27:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121319W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 13 23:17:39 2019
Response via : Initial Calibration

