

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121620\
 Data File : VN065157.D
 Acq On : 15 Dec 2020 22:07
 Operator : JC/MD
 Sample : VSTDIC005
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 16 08:12:02 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 16 08:10:37 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	374394	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	889044	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	803472	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	208508	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	31587	7.27	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	14.54%#		
35) Dibromofluoromethane	7.547	113	23435	3.92	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	7.84%#		
50) Toluene-d8	10.058	98	110739	5.04	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	10.08%#		
62) 4-Bromofluorobenzene	12.376	95	27794	3.62	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	7.24%#		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.830	85	19470	5.29	ug/l	95
3) Chloromethane	2.036	50	20014	5.41	ug/l	95
4) Vinyl Chloride	2.164	62	31259	9.07	ug/l	97
5) Bromomethane	2.525	94	25894	11.68	ug/l	97
6) Chloroethane	2.669	64	20969	11.28	ug/l	99
7) Trichlorofluoromethane	2.984	101	51610	8.38	ug/l	97
8) Diethyl Ether	3.370	74	10247	5.38	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.717	101	18453	5.45	ug/l	97
10) Methyl Iodide	3.914	142	21363	4.83	ug/l	98
11) Tert butyl alcohol	4.730	59	13539	27.14	ug/l #	71
12) 1,1-Dichloroethene	3.701	96	17530	5.44	ug/l #	79
13) Acrolein	3.566	56	9210	29.49	ug/l	95
14) Allyl chloride	4.277	41	22092	4.42	ug/l	89
15) Acrylonitrile	4.936	53	34089	23.51	ug/l	96
16) Acetone	3.775	43	33139	22.63	ug/l	97
17) Carbon Disulfide	4.013	76	47323	5.26	ug/l	97
18) Methyl Acetate	4.274	43	15464	4.65	ug/l #	82
19) Methyl tert-butyl Ether	4.984	73	48267	5.06	ug/l	98
20) Methylene Chloride	4.505	84	21428	4.89	ug/l	93
21) trans-1,2-Dichloroethene	4.987	96	17584	5.10	ug/l	99
22) Diisopropyl ether	5.901	45	50775	4.93	ug/l #	80
23) Vinyl Acetate	5.846	43	200791	23.09	ug/l	99
24) 1,1-Dichloroethane	5.801	63	34542	5.36	ug/l #	92
25) 2-Butanone	6.788	43	48728	22.77	ug/l	95
26) 2,2-Dichloropropane	6.775	77	30318	5.49	ug/l #	72
27) cis-1,2-Dichloroethene	6.782	96	20918	5.05	ug/l	91
28) Bromochloromethane	7.155	49	18882	6.30	ug/l	96
29) Tetrahydrofuran	7.167	42	38557	29.07	ug/l	95
30) Chloroform	7.325	83	47628	6.89	ug/l	97
31) Cyclohexane	7.618	56	37971	6.58	ug/l #	82
32) 1,1,1-Trichloroethane	7.528	97	40625	6.63	ug/l	96
36) 1,1-Dichloropropene	7.753	75	33335	4.09	ug/l	98
37) Ethyl Acetate	6.875	43	17179	2.34	ug/l #	92
38) Carbon Tetrachloride	7.733	117	39506	4.37	ug/l	95
39) Methylcyclohexane	9.048	83	41207	4.69	ug/l	92
40) Benzene	8.003	78	98055	4.00	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121620\
 Data File : VN065157.D
 Acq On : 15 Dec 2020 22:07
 Operator : JC/MD
 Sample : VSTDIC005
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Dec 16 08:12:02 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 16 08:10:37 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.116	41	6875	1.84	ug/l #	100
42) 1,2-Dichloroethane	8.084	62	38830	4.49	ug/l	97
43) Isopropyl Acetate	8.126	43	50367	4.10	ug/l #	85
44) Trichloroethene	8.804	130	35392	4.86	ug/l	99
45) 1,2-Dichloropropane	9.084	63	34325	5.15	ug/l	94
46) Dibromomethane	9.174	93	22291	4.96	ug/l	99
47) Bromodichloromethane	9.370	83	42169	4.66	ug/l	99
48) Methyl methacrylate	9.167	41	22781	4.01	ug/l	92
49) 1,4-Dioxane	9.164	88	9308	94.39	ug/l #	94
51) 4-Methyl-2-Pentanone	9.952	43	156501	22.43	ug/l	98
52) Toluene	10.122	92	78187	5.04	ug/l	99
53) t-1,3-Dichloropropene	10.354	75	42171	4.57	ug/l	99
54) cis-1,3-Dichloropropene	9.807	75	49113	4.84	ug/l	97
55) 1,1,2-Trichloroethane	10.534	97	31267	5.08	ug/l	98
56) Ethyl methacrylate	10.402	69	35030	4.48	ug/l	96
57) 1,3-Dichloropropane	10.679	76	51189	5.08	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.663	63	84421	21.22	ug/l	98
59) 2-Hexanone	10.724	43	106709	21.80	ug/l	90
60) Dibromochloromethane	10.875	129	34079	4.75	ug/l	94
61) 1,2-Dibromoethane	10.978	107	32417	5.05	ug/l	98
64) Tetrachloroethene	10.601	164	38201	4.96	ug/l	97
65) Chlorobenzene	11.402	112	85081	5.01	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.479	131	31833	4.75	ug/l	97
67) Ethyl Benzene	11.482	91	130547	4.62	ug/l	99
68) m/p-Xylenes	11.595	106	97303	8.84	ug/l	98
69) o-Xylene	11.920	106	46040	4.48	ug/l	99
70) Styrene	11.936	104	69283	4.14	ug/l	98
71) Bromoform	12.100	173	21367	4.02	ug/l #	99
73) Isopropylbenzene	12.222	105	104090	7.56	ug/l	99
74) N-amyl acetate	12.045	43	36544	6.99	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.476	83	25974	6.24	ug/l	98
76) 1,2,3-Trichloropropane	12.527	75	33951	8.73	ug/l #	100
77) Bromobenzene	12.498	156	20496	5.35	ug/l	95
78) n-propylbenzene	12.563	91	83311	5.35	ug/l	96
79) 2-Chlorotoluene	12.646	91	53513	5.53	ug/l	98
80) 1,3,5-Trimethylbenzene	12.704	105	63976	5.58	ug/l	95
81) trans-1,4-Dichloro-2-b...	12.273	75	8926	6.55	ug/l #	77
82) 4-Chlorotoluene	12.746	91	49812	5.08	ug/l	96
83) tert-Butylbenzene	12.968	119	52882	5.53	ug/l	95
84) 1,2,4-Trimethylbenzene	13.013	105	59472	5.15	ug/l	99
85) sec-Butylbenzene	13.145	105	70042	5.65	ug/l	97
86) p-Isopropyltoluene	13.260	119	59735	5.27	ug/l	93
87) 1,3-Dichlorobenzene	13.257	146	33537	4.90	ug/l	96
88) 1,4-Dichlorobenzene	13.338	146	34950	4.99	ug/l	92
89) n-Butylbenzene	13.588	91	50963	5.39	ug/l	93
90) Hexachloroethane	13.846	117	12159	5.53	ug/l	87
91) 1,2-Dichlorobenzene	13.627	146	33095	5.04	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.248	75	4512	5.50	ug/l	79
93) 1,2,4-Trichlorobenzene	14.884	180	8303	2.66	ug/l	88
94) Hexachlorobutadiene	14.977	225	10815	5.13	ug/l	98
95) Naphthalene	15.103	128	20205	2.46	ug/l	97
96) 1,2,3-Trichlorobenzene	15.283	180	8538	2.96	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121620\
Data File : VN065157.D
Acq On : 15 Dec 2020 22:07
Operator : JC/MD
Sample : VSTDIC005
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Dec 16 08:12:02 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121620W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 16 08:10:37 2020
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 16 08:12:02 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121620W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 16 08:10:37 2020
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:168 Resp: 374394
Ion Ratio Lower Upper
168 100
99 60.6 42.4 63.6



#2
Dichlorodifluoromethane
Concen: 5.29 ug/l
RT: 1.830 min Scan# 28
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 85 Resp: 19470
Ion Ratio Lower Upper
85 100
87 36.4 16.7 50.1









#3
Chloromethane
Concen: 5.41 ug/l
RT: 2.036 min Scan# 92
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 50 Resp: 20014
Ion Ratio Lower Upper
50 100
52 35.8 26.3 39.5



#4
Vinyl Chloride
Concen: 9.07 ug/l
RT: 2.164 min Scan# 132
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07


Tgt Ion: 62 Resp: 31259
Ion Ratio Lower Upper
62 100
64 32.3 24.6 36.8








#5
Bromomethane
Concen: 11.68 ug/l
RT: 2.525 min Scan# 244
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 94 Resp: 25894
Ion Ratio Lower Upper
94 100
96 95.8 74.0 111.0



#6
Chloroethane
Concen: 11.28 ug/l
RT: 2.669 min Scan# 289
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 64 Resp: 20969
Ion Ratio Lower Upper
64 100
66 30.4 24.8 37.2









#7
Trichlorofluoromethane
Concen: 8.38 ug/l
RT: 2.984 min Scan# 387
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:101 Resp: 51610
Ion Ratio Lower Upper
101 100
103 67.0 51.4 77.2



#8
Diethyl Ether
Concen: 5.38 ug/l
RT: 3.370 min Scan# 507
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 74 Resp: 10247
Ion Ratio Lower Upper
74 100
45 94.9 46.0 137.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.45 ug/l
RT: 3.717 min Scan# 615
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:101 Resp: 18453
Ion Ratio Lower Upper
101 100
85 42.6 37.0 55.4
151 78.3 61.1 91.7



#10
Methyl Iodide
Concen: 4.83 ug/l
RT: 3.914 min Scan# 676
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:142 Resp: 21363
Ion Ratio Lower Upper
142 100
127 51.7 40.6 61.0
141 16.4 11.9 17.9









#11
Tert butyl alcohol
Concen: 27.14 ug/l
RT: 4.730 min Scan# 930
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 59 Resp: 13539
Ion Ratio Lower Upper
59 100
57 0.0 8.8 13.2#



#12
1,1-Dichloroethene
Concen: 5.44 ug/l
RT: 3.701 min Scan# 610
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 96 Resp: 17530
Ion Ratio Lower Upper
96 100
61 138.7 131.2 196.8
98 46.2 53.5 80.3#





#13

Acrolein

Concen: 29.49 ug/l

RT: 3.566 min Scan# 568

Delta R.T. 0.000 min

Lab File: VN065157.D





Acq: 15 Dec 2020 22:07

Tgt Ion: 56 Resp: 9210

Ion Ratio Lower Upper

56 100

55 69.6 59.2 88.8



#14

Allyl chloride

Concen: 4.42 ug/l

RT: 4.277 min Scan# 789

Delta R.T. 0.003 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07




Tgt Ion: 41 Resp: 22092


Ion Ratio Lower Upper

41 100

39 95.7 66.6 99.8

76 40.6 30.0 45.0









#15
Acrylonitrile
Concen: 23.51 ug/l
RT: 4.936 min Scan# 994
Delta R.T. 0.010 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 53 Resp: 34089
Ion Ratio Lower Upper
53 100
52 79.6 67.3 100.9
51 37.1 30.2 45.2



#16
Acetone
Concen: 22.63 ug/l
RT: 3.775 min Scan# 633
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 43 Resp: 33139
Ion Ratio Lower Upper
43 100
58 28.9 21.9 32.9





#17

Carbon Disulfide


Concen: 5.26 ug/l

RT: 4.013 min Scan# 707

Delta R.T. 0.003 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07






Tgt Ion: 76 Resp: 47323

Ion Ratio Lower Upper

76 100

78 8.7 7.7 11.5



#18

Methyl Acetate


Concen: 4.65 ug/l

RT: 4.274 min Scan# 788

Delta R.T. -0.003 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07






Tgt Ion: 43 Resp: 15464

Ion Ratio Lower Upper

43 100

74 16.5 20.6 31.0#









#19
Methyl tert-butyl Ether
Concen: 5.06 ug/l
RT: 4.984 min Scan# 1009
Delta R.T. -0.006 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 73 Resp: 48267
Ion Ratio Lower Upper
73 100
57 20.2 16.9 25.3



#20
Methylene Chloride
Concen: 4.89 ug/l
RT: 4.505 min Scan# 860
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 84 Resp: 21428
Ion Ratio Lower Upper
84 100
49 117.7 100.2 150.4
51 29.6 29.5 44.3
86 60.6 52.4 78.6









#21
trans-1,2-Dichloroethene
Concen: 5.10 ug/l
RT: 4.987 min Scan# 1010
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 17584
Ion Ratio Lower Upper
96 100
61 136.4 111.1 166.7
98 62.6 50.2 75.2



#22
Diisopropyl ether
Concen: 4.93 ug/l
RT: 5.901 min Scan# 1294
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 45 Resp: 50775
Ion Ratio Lower Upper
45 100
43 86.3 50.3 75.5#
87 28.0 23.8 35.8
59 13.5 9.4 14.2









#23
 Vinyl Acetate
 Concen: 23.09 ug/l
 RT: 5.846 min Scan# 1277
 Delta R.T. 0.006 min
 Lab File: VN065157.D
 Acq: 15 Dec 2020 22:07




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion: 43 Resp: 200791
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 5.36 ug/l
 RT: 5.801 min Scan# 1263
 Delta R.T. 0.010 min
 Lab File: VN065157.D
 Acq: 15 Dec 2020 22:07

Tgt Ion: 63 Resp: 34542
 Ion Ratio Lower Upper
 63 100
 98 9.7 2.9 8.7#
 100 4.9 2.1 6.2









#25
2-Butanone
Concen: 22.77 ug/l
RT: 6.788 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 48728
Ion Ratio Lower Upper
43 100
72 26.4 23.1 34.7



#26
2,2-Dichloropropane
Concen: 5.49 ug/l
RT: 6.775 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 77 Resp: 30318
Ion Ratio Lower Upper
77 100
97 10.1 11.9 35.9#









#27
 cis-1,2-Dichloroethene
 Concen: 5.05 ug/l
 RT: 6.782 min Scan# 1568
 Delta R.T. 0.000 min
 Lab File: VN065157.D
 Acq: 15 Dec 2020 22:07




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 96 Resp: 20918
 Ion Ratio Lower Upper
 96 100
 61 148.0 0.0 267.8
 98 65.2 0.0 128.0



#28
 Bromochloromethane
 Concen: 6.30 ug/l
 RT: 7.155 min Scan# 1684
 Delta R.T. 0.003 min
 Lab File: VN065157.D
 Acq: 15 Dec 2020 22:07

Tgt Ion: 49 Resp: 18882
 Ion Ratio Lower Upper
 49 100
 129 0.2 0.0 2.2
 130 84.8 71.2 106.8





#29

Tetrahydrofuran

Concen: 29.07 ug/l





RT: 7.167 min Scan# 1688

Delta R.T. 0.003 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07

Tgt Ion:	42	Resp:	38557
Ion	Ratio	Lower	Upper
42	100		
72	45.6	39.0	58.4
71	42.3	37.0	55.4



#30

Chloroform

Concen: 6.89 ug/l




RT: 7.325 min Scan# 1737


Delta R.T. -0.003 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07


Tgt Ion:	83	Resp:	47628
Ion	Ratio	Lower	Upper
83	100		
85	68.1	52.6	78.8








#31
Cyclohexane
Concen: 6.58 ug/l
RT: 7.618 min Scan# 1828
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 56 Resp: 37971
Ion Ratio Lower Upper
56 100
69 71.2 28.1 42.1#
84 91.4 72.0 108.0



#32
1,1,1-Trichloroethane
Concen: 6.63 ug/l
RT: 7.528 min Scan# 1800
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 97 Resp: 40625
Ion Ratio Lower Upper
97 100
99 64.4 50.5 75.7
61 45.1 32.6 49.0









#33
1,2-Dichloroethane-d4
Concen: 7.27 ug/l
RT: 7.984 min Scan# 1942
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 65 Resp: 31587
Ion Ratio Lower Upper
65 100
67 54.1 0.0 111.0



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.547 min Scan# 2117
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 114 Resp: 889044
Ion Ratio Lower Upper
114 100
63 19.0 0.0 37.6
88 15.3 0.0 30.2





#35

Dibromofluoromethane

Concen: 3.92 ug/l

RT: 7.547 min Scan# 1806





Delta R.T. -0.003 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07

Tgt Ion:	113	Resp:	23435
Ion	Ratio	Lower	Upper
113	100		
111	103.5	81.6	122.4
192	21.0	15.9	23.9

Instrument :
MSVOA_N
ClientSampled :



#36

1,1-Dichloropropene

Concen: 4.09 ug/l




RT: 7.753 min Scan# 1870


Delta R.T. 0.000 min

Lab File: VN065157.D

Acq: 15 Dec 2020 22:07


Tgt Ion:	75	Resp:	33335
Ion	Ratio	Lower	Upper
75	100		
110	36.6	18.6	55.6
77	29.2	24.6	36.8








#37
Ethyl Acetate
Concen: 2.34 ug/l
RT: 6.875 min Scan# 1597
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion: 43 Resp: 17179
Ion Ratio Lower Upper
43 100
61 11.2 12.9 19.3#
70 12.4 9.0 13.6



#38
Carbon Tetrachloride
Concen: 4.37 ug/l
RT: 7.733 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 117 Resp: 39506
Ion Ratio Lower Upper
117 100
119 100.6 75.9 113.9
121 30.6 24.3 36.5









#39
Methylcyclohexane
Concen: 4.69 ug/l
RT: 9.048 min Scan# 2273
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 83 Resp: 41207
Ion Ratio Lower Upper
83 100
55 82.6 57.4 86.0
98 44.3 35.8 53.8



#40
Benzene
Concen: 4.00 ug/l
RT: 8.003 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07


Tgt Ion: 78 Resp: 98055
Ion Ratio Lower Upper
78 100
77 23.8 18.5 27.7








#41
Methacrylonitrile
Concen: 1.84 ug/l
RT: 7.116 min Scan# 1672
Delta R.T. -0.022 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Instrument :
MSVOA_N
ClientSampled :







Tgt Ion	Ratio	Lower	Upper
41	100		
39	86.6	0.0	0.0#
67	147.4	0.0	0.0#
52	35.8	0.0	0.0#



#42
1,2-Dichloroethane
Concen: 4.49 ug/l
RT: 8.084 min Scan# 1973
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion	Ratio	Lower	Upper
62	100		
98	9.0	0.0	20.4









#43

Isopropyl Acetate
Concen: 4.10 ug/l
RT: 8.126 min Scan# 1986
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Instrument :
MSVOA_N
ClientSampleId :




Tgt Ion: 43 Resp: 50367
Ion Ratio Lower Upper
43 100
61 22.4 24.0 36.0#
87 7.7 12.2 18.4#




#44

Trichloroethene
Concen: 4.86 ug/l
RT: 8.804 min Scan# 2197
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:130 Resp: 35392
Ion Ratio Lower Upper
130 100
95 96.6 0.0 194.4









#45
1,2-Dichloropropane
Concen: 5.15 ug/l
RT: 9.084 min Scan# 2284
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 63 Resp: 34325
Ion Ratio Lower Upper
63 100
65 28.0 25.1 37.7



#46
Dibromomethane
Concen: 4.96 ug/l
RT: 9.174 min Scan# 2312
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07


Tgt Ion: 93 Resp: 22291
Ion Ratio Lower Upper
93 100
95 84.2 66.6 99.8
174 101.6 82.5 123.7








#47
Bromodichloromethane
Concen: 4.66 ug/l
RT: 9.370 min Scan# 2373
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07


Instrument :
MSVOA_N
ClientSampled :





Tgt Ion: 83 Resp: 42169
Ion Ratio Lower Upper
83 100
85 64.5 51.4 77.0
127 7.9 7.4 11.0




#48
Methyl methacrylate
Concen: 4.01 ug/l
RT: 9.167 min Scan# 2310
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07



Tgt Ion: 41 Resp: 22781
Ion Ratio Lower Upper
41 100
69 94.2 79.8 119.6
39 60.0 40.8 61.2









#49
1,4-Dioxane
Concen: 94.39 ug/l
RT: 9.164 min Scan# 2309
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 88 Resp: 9308
Ion Ratio Lower Upper
88 100
43 34.1 21.5 32.3#
58 65.7 50.6 75.8



#50
Toluene-d8
Concen: 5.04 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 98 Resp: 110739
Ion Ratio Lower Upper
98 100
100 64.6 51.8 77.6









#51
4-Methyl-2-Pentanone
Concen: 22.43 ug/l
RT: 9.952 min Scan# 2554
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 156501
Ion Ratio Lower Upper
43 100
58 41.5 34.4 51.6



#52
Toluene
Concen: 5.04 ug/l
RT: 10.122 min Scan# 2607
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 92 Resp: 78187
Ion Ratio Lower Upper
92 100
91 170.6 137.6 206.4









#53
t-1,3-Dichloropropene
Concen: 4.57 ug/l
RT: 10.354 min Scan# 2679
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 42171
Ion Ratio Lower Upper
75 100
77 30.9 25.3 37.9



#54
cis-1,3-Dichloropropene
Concen: 4.84 ug/l
RT: 9.807 min Scan# 2509
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 75 Resp: 49113
Ion Ratio Lower Upper
75 100
77 29.8 25.7 38.5
39 40.2 31.3 46.9









#55
1,1,2-Trichloroethane
Concen: 5.08 ug/l
RT: 10.534 min Scan# 2735
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	97	Resp:	31267
Ion	Ratio	Lower	Upper
97	100		
83	83.8	68.0	102.0
85	53.4	44.7	67.1
99	61.1	48.2	72.2



#56
Ethyl methacrylate
Concen: 4.48 ug/l
RT: 10.402 min Scan# 2694
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:	69	Resp:	35030
Ion	Ratio	Lower	Upper
69	100		
41	58.3	45.6	68.4
39	33.3	23.0	34.4









#57
1,3-Dichloropropane
Concen: 5.08 ug/l
RT: 10.679 min Scan# 2780
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 76 Resp: 51189
Ion Ratio Lower Upper
76 100
78 33.5 26.2 39.4



#58
2-Chloroethyl Vinyl ether
Concen: 21.22 ug/l
RT: 9.663 min Scan# 2464
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 63 Resp: 84421
Ion Ratio Lower Upper
63 100
106 29.0 22.5 33.7









#59
2-Hexanone
Concen: 21.80 ug/l
RT: 10.724 min Scan# 2794
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 43 Resp: 106709
Ion Ratio Lower Upper
43 100
58 52.1 29.6 88.9



#60
Dibromochloromethane
Concen: 4.75 ug/l
RT: 10.875 min Scan# 2841
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 129 Resp: 34079
Ion Ratio Lower Upper
129 100
127 81.3 38.1 114.5









#61
1,2-Dibromoethane
Concen: 5.05 ug/l
RT: 10.978 min Scan# 2873
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:107 Resp: 32417
Ion Ratio Lower Upper
107 100
109 91.6 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 3.62 ug/l
RT: 12.376 min Scan# 3308
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 95 Resp: 27794
Ion Ratio Lower Upper
95 100
174 72.2 0.0 162.8
176 73.6 0.0 160.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.380 min Scan# 2998
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:117 Resp: 803472
Ion Ratio Lower Upper
117 100
82 53.8 43.8 65.6
119 31.4 26.0 39.0



#64
Tetrachloroethene
Concen: 4.96 ug/l
RT: 10.601 min Scan# 2756
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:164 Resp: 38201
Ion Ratio Lower Upper
164 100
166 123.9 103.0 154.6
129 95.8 75.4 113.2
131 94.0 72.8 109.2









#65
Chlorobenzene
Concen: 5.01 ug/l
RT: 11.402 min Scan# 3005
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 85081
Ion Ratio Lower Upper
112 100
114 28.8 25.5 38.3



#66
1,1,1,2-Tetrachloroethane
Concen: 4.75 ug/l
RT: 11.479 min Scan# 3029
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:131 Resp: 31833
Ion Ratio Lower Upper
131 100
133 95.6 48.1 144.4
119 71.9 33.8 101.4









#67
Ethyl Benzene
Concen: 4.62 ug/l
RT: 11.482 min Scan# 3030
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 130547
Ion Ratio Lower Upper
91 100
106 31.5 25.6 38.4



#68
m/p-Xylenes
Concen: 8.84 ug/l
RT: 11.595 min Scan# 3065
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:106 Resp: 97303
Ion Ratio Lower Upper
106 100
91 201.3 158.4 237.6









#69
o-Xylene
Concen: 4.48 ug/l
RT: 11.920 min Scan# 3166
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:106 Resp: 46040
Ion Ratio Lower Upper
106 100
91 210.2 104.4 313.2



#70
Styrene
Concen: 4.14 ug/l
RT: 11.936 min Scan# 3171
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:104 Resp: 69283
Ion Ratio Lower Upper
104 100
78 49.6 38.7 58.1
103 55.8 42.9 64.3









#71
Bromoform
Concen: 4.02 ug/l
RT: 12.100 min Scan# 3222
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 21367
Ion Ratio Lower Upper
173 100
175 48.7 24.1 72.2
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.315 min Scan# 3600
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:152 Resp: 208508
Ion Ratio Lower Upper
152 100
115 59.9 28.5 85.5
150 154.3 0.0 349.6









#73
Isopropylbenzene
Concen: 7.56 ug/l
RT: 12.222 min Scan# 3260
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:105 Resp: 104090
Ion Ratio Lower Upper
105 100
120 26.4 13.5 40.4



#74
N-ethyl acetate
Concen: 6.99 ug/l
RT: 12.045 min Scan# 3205
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 43 Resp: 36544
Ion Ratio Lower Upper
43 100
70 40.2 37.8 56.6
55 26.9 21.9 32.9
61 22.2 21.4 32.2









#75
1,1,2,2-Tetrachloroethane
Concen: 6.24 ug/l
RT: 12.476 min Scan# 3339
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 25974
Ion Ratio Lower Upper
83 100
131 10.8 5.1 15.4
85 65.9 32.1 96.3



#76
1,2,3-Trichloropropane
Concen: 8.73 ug/l
RT: 12.527 min Scan# 3355
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 75 Resp: 33951
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 5.35 ug/l
RT: 12.498 min Scan# 3346
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:156 Resp: 20496
Ion Ratio Lower Upper
156 100
77 204.0 97.9 293.7
158 93.3 48.7 146.1



#78
n-propylbenzene
Concen: 5.35 ug/l
RT: 12.563 min Scan# 3366
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 91 Resp: 83311
Ion Ratio Lower Upper
91 100
120 21.5 11.8 35.4









#79
2-Chlorotoluene
Concen: 5.53 ug/l
RT: 12.646 min Scan# 3392
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91 Resp: 53513
Ion Ratio Lower Upper
91 100
126 33.7 17.5 52.6



#80
1,3,5-Trimethylbenzene
Concen: 5.58 ug/l
RT: 12.704 min Scan# 3410
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 105 Resp: 63976
Ion Ratio Lower Upper
105 100
120 45.7 24.6 73.8









#81
trans-1,4-Dichloro-2-butene
Concen: 6.55 ug/l
RT: 12.273 min Scan# 3276
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 75 Resp: 8926
Ion Ratio Lower Upper
75 100
53 113.6 67.1 100.7#
89 50.4 37.2 55.8



#82
4-Chlorotoluene
Concen: 5.08 ug/l
RT: 12.746 min Scan# 3423
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 91 Resp: 49812
Ion Ratio Lower Upper
91 100
126 36.8 17.2 51.5









#83
tert-Butylbenzene
Concen: 5.53 ug/l
RT: 12.968 min Scan# 3492
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 52882
Ion Ratio Lower Upper
119 100
91 67.9 31.3 93.9
134 24.8 12.6 37.8



#84
1,2,4-Trimethylbenzene
Concen: 5.15 ug/l
RT: 13.013 min Scan# 3506
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:105 Resp: 59472
Ion Ratio Lower Upper
105 100
120 46.4 22.9 68.5









#85
 sec-Butylbenzene
 Concen: 5.65 ug/l
 RT: 13.145 min Scan# 3547
 Delta R.T. 0.000 min
 Lab File: VN065157.D
 Acq: 15 Dec 2020 22:07




Instrument :
 MSVOA_N
 ClientSampled :


Tgt Ion:105 Resp: 70042
 Ion Ratio Lower Upper
 105 100
 134 18.4 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 5.27 ug/l
 RT: 13.260 min Scan# 3583
 Delta R.T. 0.000 min
 Lab File: VN065157.D
 Acq: 15 Dec 2020 22:07

Tgt Ion:119 Resp: 59735
 Ion Ratio Lower Upper
 119 100
 134 25.0 13.3 39.8
 91 28.5 11.5 34.5









#87
1,3-Dichlorobenzene
Concen: 4.90 ug/l
RT: 13.257 min Scan# 3582
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:146 Resp: 33537
Ion Ratio Lower Upper
146 100
111 44.6 19.7 59.1
148 64.8 32.1 96.3



#88
1,4-Dichlorobenzene
Concen: 4.99 ug/l
RT: 13.338 min Scan# 3607
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:146 Resp: 34950
Ion Ratio Lower Upper
146 100
111 44.6 19.4 58.2
148 68.9 31.6 95.0









#89
n-Butylbenzene
Concen: 5.39 ug/l
RT: 13.588 min Scan# 3685
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion: 91	Resp: 50963
Ion Ratio	Lower Upper
91 100	
92 45.6	25.5 76.5
134 23.2	13.2 39.5



#90
Hexachloroethane
Concen: 5.53 ug/l
RT: 13.846 min Scan# 3765
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion: 117	Resp: 12159
Ion Ratio	Lower Upper
117 100	
201 71.1	41.2 123.6









#91
1,2-Dichlorobenzene
Concen: 5.04 ug/l
RT: 13.627 min Scan# 3697
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	146	Resp:	33095
Ion Ratio	Lower	Upper	
146	100		
111	42.1	20.5	61.5
148	64.0	31.9	95.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.50 ug/l
RT: 14.248 min Scan# 3890
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:	75	Resp:	4512
Ion Ratio	Lower	Upper	
75	100		
155	61.5	43.5	130.4
157	95.9	56.0	168.0









#93
1,2,4-Trichlorobenzene
Concen: 2.66 ug/l
RT: 14.884 min Scan# 4088
Delta R.T. 0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07




Instrument :
MSVOA_N
ClientSampled :


Tgt Ion:180 Resp: 8303
Ion Ratio Lower Upper
180 100
182 105.2 47.0 141.2
145 38.2 15.3 45.8



#94
Hexachlorobutadiene
Concen: 5.13 ug/l
RT: 14.977 min Scan# 4117
Delta R.T. -0.003 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:225 Resp: 10815
Ion Ratio Lower Upper
225 100
223 60.1 31.2 93.6
227 61.4 30.9 92.8









#95
Naphthalene
Concen: 2.46 ug/l
RT: 15.103 min Scan# 4156
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Instrument :
MSVOA_N
ClientSampled :

Tgt Ion:128 Resp: 20205
Ion Ratio Lower Upper
128 100
127 15.2 10.2 15.2
129 10.9 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 2.96 ug/l
RT: 15.283 min Scan# 4212
Delta R.T. 0.000 min
Lab File: VN065157.D
Acq: 15 Dec 2020 22:07

Tgt Ion:180 Resp: 8538
Ion Ratio Lower Upper
180 100
182 93.2 47.6 142.9
145 35.2 16.2 48.5

