

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121620\
 Data File : VN065171.D
 Acq On : 16 Dec 2020 12:57
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Dec 21 13:58:52 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121520w.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 15 15:18:54 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	104	0.00
2 M	Dichlorodifluoromethane	1.701	1.700	0.1	112	0.00
3 M	Chloromethane	1.959	1.971	-0.6	110	0.00
4 M	Vinyl Chloride	2.158	1.933	10.4	106	0.00
5 M	Bromomethane	1.515	1.238	18.3	111	0.00
6 M	Chloroethane	1.286	1.076	16.3	110	0.00
7 M	Trichlorofluoromethane	3.193	2.786	12.7	110	0.00
8 T	Diethyl Ether	1.020	1.015	0.5	113	0.00
9	1,1,2-Trichlorotrifluoroeth	1.648	1.625	1.4	112	0.00
10 M	1,1-Dichloroethene	1.599	1.575	1.5	112	0.00
11	Methyl Iodide	2.223	2.165	2.6	113	0.00
12	Methyl Acetate	1.455	1.385	4.8	108	0.00
13 M	Acrolein	0.206	0.179	13.1	101	0.00
14 M	Acrylonitrile	0.713	0.689	3.4	108	0.00
15 M	Acetone	0.224	0.202	9.8	91	0.00
16 M	Carbon Disulfide	4.704	4.598	2.3	110	0.00
17	Allyl chloride	2.316	2.288	1.2	124	0.00
18 M	Methylene Chloride	1.881	1.847	1.8	108	0.00
19 M	trans-1,2-Dichloroethene	1.737	1.705	1.8	111	0.00
20 T	Diisopropyl ether	5.713	5.112	10.5	112	0.00
21 M	1,1-Dichloroethane	3.447	3.085	10.5	110	0.00
22 M	cis-1,2-Dichloroethene	1.958	1.908	2.6	110	0.00
23 M	tert-Butyl Alcohol	0.231	0.212	8.2	111	0.00
24 M	Methyl tert-Butyl Ether	4.844	4.711	2.7	113	0.00
25 M	Chloroform	3.258	3.166	2.8	110	0.00
26	Cyclohexane	2.519	2.450	2.7	110	0.00
27 s	1,2-Dichloroethane-d4	2.054	1.982	3.5	105	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
29	1,1-Dichloropropene	0.472	0.461	2.3	112	0.00
30 M	2-Butanone	0.183	0.169	7.7	103	0.00
31	2,2-Dichloropropane	0.523	0.498	4.8	111	0.00
32 M	1,1,1-Trichloroethane	0.561	0.531	5.3	110	0.00
33 M	Carbon Tetrachloride	0.507	0.475	6.3	108	0.00
34 M	Benzene	1.427	1.421	0.4	110	0.00
35	Methacrylonitrile	0.179	0.127	29.1	76	0.00
36 M	1,2-Dichloroethane	0.489	0.463	5.3	110	0.00
37 M	Trichloroethene	0.389	0.383	1.5	111	0.00
38	Methylcyclohexane	0.563	0.507	9.9	113	0.00
39 M	1,2-Dichloropropane	0.383	0.368	3.9	110	0.00
40	Dibromomethane	0.269	0.240	10.8	109	0.00
41 M	Bromodichloromethane	0.546	0.480	12.1	110	0.00
42 M	Vinyl Acetate	0.871	0.763	12.4	110	0.00
43	Ethyl Acetate	0.377	0.329	12.7	95	0.00
44	Isopropyl Acetate	0.630	0.580	7.9	107	0.00
45 T	1,4-Dioxane	0.006	0.005#	16.7	104	0.00
46	Methyl methacrylate	0.320	0.262	18.1	105	0.00
47	n-aryl Acetate	0.521	0.464	10.9	110	0.00
48 M	t-1,3-Dichloropropene	0.534	0.502	6.0	112	0.00
49 T	cis-1,3-Dichloropropene	0.606	0.562	7.3	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121620\
 Data File : VN065171.D
 Acq On : 16 Dec 2020 12:57
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Dec 21 13:58:52 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N121520w.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Dec 15 15:18:54 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 M	1,1,2-Trichloroethane	0.349	0.350	-0.3	112	0.00
51	Ethyl methacrylate	0.467	0.423	9.4	107	0.00
52	1,3-Dichloropropane	0.578	0.575	0.5	113	0.00
53 M	Dibromochloromethane	0.404	0.385	4.7	111	0.00
54 M	1,2-Dibromoethane	0.364	0.348	4.4	110	0.00
55 M	2-Chloroethyl vinyl ether	0.236	0.212	10.2	107	0.00
56 M	Bromoform	0.275	0.254	7.6	106	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
58 M	4-Methyl-2-Pentanone	0.437	0.382	12.6	107	0.00
59 M	2-Hexanone	0.290	0.261	10.0	105	0.00
60 S	4-Bromofluorobenzene	0.471	0.466	1.1	109	0.00
61 M	Tetrachloroethene	0.436	0.453	-3.9	114	0.00
62 M	Toluene	1.662	1.619	2.6	110	0.00
63 S	Toluene-d8	1.448	1.400	3.3	105	0.00
64 M	Chlorobenzene	1.036	1.021	1.4	113	0.00
65	1,1,1,2-Tetrachloroethane	0.393	0.374	4.8	109	0.00
66 M	Ethyl Benzene	1.761	1.692	3.9	112	0.00
67 M	m/p-Xylenes	0.686	0.672	2.0	114	0.00
68 M	o-Xylene	0.649	0.623	4.0	112	0.00
69 M	Styrene	1.079	1.028	4.7	113	0.00
70	Isopropylbenzene	1.694	1.664	1.8	113	0.00
71 M	1,1,2,2-Tetrachloroethane	0.491	0.459	6.5	107	0.00
72	1,2,3-Trichloropropane	0.470	0.449	4.5	104	0.00
73	Bromobenzene	0.440	0.440	0.0	116	0.00
74	n-propylbenzene	1.933	1.872	3.2	113	0.00
75	2-Chlorotoluene	1.176	1.151	2.1	113	0.00
76	1,3,5-Trimethylbenzene	1.462	1.409	3.6	111	0.00
77	t-1,4-Dichloro-2-butene	0.166	0.147	11.4	107	0.00
78	4-Chlorotoluene	1.201	1.141	5.0	112	0.00
79	tert-butylbenzene	1.227	1.176	4.2	113	0.00
80	1,2,4-Trimethylbenzene	1.450	1.418	2.2	113	0.00
81	sec-Butylbenzene	1.592	1.545	3.0	115	0.00
82	p-Isopropyltoluene	1.496	1.411	5.7	117	0.00
83 M	1,3-Dichlorobenzene	0.784	0.719	8.3	112	0.00
84 M	1,4-Dichlorobenzene	0.821	0.721	12.2	116	0.00
85	n-Butylbenzene	1.320	1.149	13.0	118	0.00
86 T	Hexachloroethane	0.288	0.240	16.7	112	0.00
87 M	1,2-Dichlorobenzene	0.813	0.712	12.4	112	0.00
88	1,2-Dibromo-3-Chloropropane	0.096	0.080	16.7	113	0.00
89	1,2,4-Trichlorobenzene	0.366	0.295	19.4	116	0.00
90	Hexachlorobutadiene	0.231	0.211	8.7	122	0.00
91 M	Naphthalene	1.033	0.729	29.4	106	0.00
92	1,2,3-Trichlorobenzene	0.345	0.269	22.0	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0