

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121621\
 Data File : VN070145.D
 Acq On : 16 Dec 2021 18:23
 Operator : JC/MD
 Sample : M5073-04ME
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-03-121321ME

Manual Integrations
 APPROVED

Reviewed By : John Carlone 12/17/2021
 Supervised By : Mahesh Dadoda 12/23/2021

Quant Time: Dec 17 01:02:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1156569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2203185	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2420933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1166193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	930331	52.098	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	104.200%	
35) Dibromofluoromethane	8.019	113	602226	43.639	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	87.280%	
50) Toluene-d8	10.440	98	2817740	50.850	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	101.700%	
62) 4-Bromofluorobenzene	12.733	95	1275721	63.548	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	127.100%	
Target Compounds						Qvalue
11) Tert butyl alcohol	5.307	59	70673	20.937	ug/l #	72
16) Acetone	4.309	43	17887m	1.665	ug/l	
20) Methylene Chloride	5.055	84	9405	0.659	ug/l	95
25) 2-Butanone	7.353	43	160100	10.969	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121621\
Data File : VN070145.D
Acq On : 16 Dec 2021 18:23
Operator : JC/MD
Sample : M5073-04ME
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-03-121321ME

Quant Time: Dec 17 01:02:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/17/2021
Supervised By : Mahesh Dadoda 12/23/2021

