

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121621\
 Data File : VN070151.D
 Acq On : 16 Dec 2021 20:56
 Operator : JC/MD
 Sample : M5131-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 211215004-02-TRIP-BLANK

Quant Time: Dec 17 01:05:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	877216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	1483548	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1317911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	467520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	748342	55.252	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.500%
35) Dibromofluoromethane	8.024	113	475431	51.162	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.320%
50) Toluene-d8	10.443	98	1864703	49.975	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.734	95	608837	45.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.080%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	33936	4.165	ug/l	96
25) 2-Butanone	7.348	43	40292	3.640	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121621\
Data File : VN070151.D
Acq On : 16 Dec 2021 20:56
Operator : JC/MD
Sample : M5131-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
211215004-02-TRIP-BLANK

Quant Time: Dec 17 01:05:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

