


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121621\
Data File : VN070126.D
Acq On : 16 Dec 2021 9:46
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Title : SW846 8260
Last Update : Fri Dec 03 08:46:47 2021



AutoFind: Scans 4004, 4005, 4006; Background Corrected with Scan 3986

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.5	50883	PASS
75	95	30	60	54.5	123197	PASS
95	95	100	100	100.0	225899	PASS
96	95	5	9	6.6	14929	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.5	168405	PASS
175	174	5	9	7.3	12331	PASS
176	174	95	101	96.0	161643	PASS
177	176	5	9	6.9	11161	PASS