

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
 Data File : VN075836.D
 Acq On : 16 Dec 2022 10:28
 Operator : JC\MD
 Sample : VN1216MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1216MBL01

Quant Time: Dec 17 08:35:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	321665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	503997	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	439227	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	182497	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	166005	47.438	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.880%
35) Dibromofluoromethane	8.177	113	151315	48.480	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.960%
50) Toluene-d8	10.571	98	582608	49.925	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.840%
62) 4-Bromofluorobenzene	12.853	95	180484	46.695	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
Data File : VN075836.D
Acq On : 16 Dec 2022 10:28
Operator : JC\MD
Sample : VN1216MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1216MBL01

Quant Time: Dec 17 08:35:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

