

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
 Data File : VN075844.D
 Acq On : 16 Dec 2022 13:50
 Operator : JC\MD
 Sample : N6088-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-12-12

Quant Time: Dec 17 08:39:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	387149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	604601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	509529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	205116	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	198869	47.217	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.440%
35) Dibromofluoromethane	8.177	113	178929	47.788	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.580%
50) Toluene-d8	10.571	98	688995	49.217	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.440%
62) 4-Bromofluorobenzene	12.853	95	208630	44.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
Data File : VN075844.D
Acq On : 16 Dec 2022 13:50
Operator : JC\MD
Sample : N6088-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-12-12

Quant Time: Dec 17 08:39:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

