

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
 Data File : VN075854.D
 Acq On : 16 Dec 2022 17:50
 Operator : JC\MD
 Sample : N6110-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-SAMPLE-TANK-10-13-14-15-17

Quant Time: Dec 17 08:42:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	365045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	576132	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	494952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	204477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	188424	47.446	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.900%
35) Dibromofluoromethane	8.177	113	168877	47.332	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	10.571	98	657374	49.278	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.560%
62) 4-Bromofluorobenzene	12.853	95	201860	45.687	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.380%
Target Compounds						
16) Acetone	4.459	43	4895	3.576	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
Data File : VN075854.D
Acq On : 16 Dec 2022 17:50
Operator : JC\MD
Sample : N6110-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-SAMPLE-TANK-10-13-14-15-17

Quant Time: Dec 17 08:42:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

