

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
 Data File : VN075859.D
 Acq On : 16 Dec 2022 19:50
 Operator : JC\MD
 Sample : N6102-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Dec 17 08:44:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 14 13:03:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	364055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	555249	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	475036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	203269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	182648	46.116	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.240%
35) Dibromofluoromethane	8.177	113	166971	48.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.120%
50) Toluene-d8	10.571	98	637616	49.595	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.180%
62) 4-Bromofluorobenzene	12.853	95	198140	46.531	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.060%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	2978	2.181	ug/l #	73
19) Methyl tert-butyl Ether	5.812	73	5731	0.539	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121622\
Data File : VN075859.D
Acq On : 16 Dec 2022 19:50
Operator : JC\MD
Sample : N6102-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Dec 17 08:44:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N121422W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 14 13:03:41 2022
Response via : Initial Calibration

