



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121624\
Data File : VN085206.D
Acq On : 16 Dec 2024 08:58
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Dec 06 22:50:42 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	4870	PASS
75	95	30	60	52.5	13523	PASS
95	95	100	100	100.0	25765	PASS
96	95	5	9	7.9	2038	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	78.8	20296	PASS
175	174	5	9	8.0	1627	PASS
176	174	95	101	97.1	19712	PASS
177	176	5	9	6.9	1363	PASS