

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121625\
 Data File : VN088529.D
 Acq On : 16 Dec 2025 15:27
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Dec 17 01:43:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	307100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	698293	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	713140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	343043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	351638	60.783	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 121.560%		
35) Dibromofluoromethane	8.147	113	257117	53.133	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.260%		
50) Toluene-d8	10.541	98	921453	51.177	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.360%		
62) 4-Bromofluorobenzene	12.823	95	358201	57.980	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 115.960%		
Target Compounds						
52) Toluene	10.606	92	57372	4.533	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121625\
Data File : VN088529.D
Acq On : 16 Dec 2025 15:27
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Dec 17 01:43:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

