

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121625\
 Data File : VN088531.D
 Acq On : 16 Dec 2025 16:07
 Operator : JC\MD
 Sample : Q3852-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TK 9&10 12"

Quant Time: Dec 17 01:44:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	300923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	700237	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	708093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	340309	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	350784	61.880	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 123.760%	
35) Dibromofluoromethane	8.147	113	262505	54.096	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.200%	
50) Toluene-d8	10.541	98	922343	51.084	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.160%	
62) 4-Bromofluorobenzene	12.829	95	355347	57.359	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 114.720%	
Target Compounds						
16) Acetone	4.418	43	10804	5.270	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121625\
Data File : VN088531.D
Acq On : 16 Dec 2025 16:07
Operator : JC\MD
Sample : Q3852-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TK 9&10 12"

Quant Time: Dec 17 01:44:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

