

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121625\
 Data File : VN088533.D
 Acq On : 16 Dec 2025 16:48
 Operator : JC\MD
 Sample : Q3852-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AI COVE 16 20"

Quant Time: Dec 17 01:45:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	242771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.077	114	554593	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	557128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	258862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	266901	58.360	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.720%			
35) Dibromofluoromethane	8.147	113	206758	53.797	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.600%			
50) Toluene-d8	10.541	98	731622	51.162	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.320%			
62) 4-Bromofluorobenzene	12.829	95	269891	55.005	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 110.020%			
Target Compounds						
16) Acetone	4.418	43	11169	6.753	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121625\
Data File : VN088533.D
Acq On : 16 Dec 2025 16:48
Operator : JC\MD
Sample : Q3852-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AI COVE 16 20"

Quant Time: Dec 17 01:45:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

