

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
 Data File : VN052923.D
 Acq On : 17 Dec 2018 13:24
 Operator : MD\SY
 Sample : J6438-08 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 23688

Quant Time: Dec 18 05:50:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1321867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2038448	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1811141	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	695516	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	665248	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
35) Dibromofluoromethane	7.59	113	525612	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
50) Toluene-d8	10.09	98	2522301	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	12.40	95	808547	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
Target Compounds						
16) Acetone	3.82	43	6234	1.975	ug/l	90
25) 2-Butanone	6.84	43	16712	3.412	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
Data File : VN052923.D
Acq On : 17 Dec 2018 13:24
Operator : MD\SY
Sample : J6438-08 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
23688

Quant Time: Dec 18 05:50:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

