

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
 Data File : VN052924.D
 Acq On : 17 Dec 2018 13:51
 Operator : MD\SY
 Sample : J6438-12 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 23691

Quant Time: Dec 18 05:48:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1391895	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	2160525	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1891389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	720847	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	702893	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
35) Dibromofluoromethane	7.59	113	554136	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
50) Toluene-d8	10.09	98	2639073	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.40	95	838613	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
Target Compounds						
25) 2-Butanone	6.84	43	6212	1.204	ug/l	94
95) Naphthalene	15.13	128	6583	0.250	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121718\
Data File : VN052924.D
Acq On : 17 Dec 2018 13:51
Operator : MD\SY
Sample : J6438-12 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
23691

Quant Time: Dec 18 05:48:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

